



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**  
By Carol Day at 11:19 am, Jul 21, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940230	NAME OF AGENCY Canton Police Department	DATE OF INSPECTION 07/08/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 108 North Fifth Street, Canton		TIME OF INSPECTION 10:56 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/08/14 22:56
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories Inc LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 93752 EXP. DATE 05/15/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.100</u>	TEST 2 <u>.100</u>	TEST 3 <u>.099</u>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Datamaster 940230 complies with the Department of Health and Senior Services rules and regulations.

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Don L Cibert
---------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220328 10/02/2014	TELEPHONE NUMBER (573) 655-4099
--	------------------------------------

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office**  
2875 James Blvd.  
Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

690 NORTH 87th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CANTON POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 940230  
07/09/14  
22:56

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefg hijklmnop  
qrstuvwxy z{|}~

Operator Signature 

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CANTON POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 940230  
07/09/14

TESTING OFFICER:  
CIBERT/DONAL  
OFFICER I.D.: 112  
PERMIT NUMBER: 220328  
EXPIRATION DATE: 10/02/14  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:59
INTERNAL STANDARD	VERIFIED	22:59
EXTERNAL STANDARD	.100	23:00
BLANK TEST	.000	23:01
EXTERNAL STANDARD	.100	23:02
BLANK TEST	.000	23:03
EXTERNAL STANDARD	.099	23:04
BLANK TEST	.000	23:05

N = 3  
SIM. = .1  
AVG. = .0996

Operator Signature 

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CANTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948238  
07/08/14

ARREST TIME: 22:00  
SUBJECT NAME:  
RFI/TEST  
DOB: 11/11/11 SEX: M  
STATE/D.L.: MO/111111111  
ARRESTING OFFICER:  
N/A  
OFFICER I.D.: N/A  
TESTING OFFICER:  
CIBERT/DON/L  
OFFICER I.D.: 112  
PERMIT NUMBER: 228328  
EXPIRATION DATE: 10/02/14  
MISCELLANEOUS DATA:  
RFI TEST  
MONTHLY MAINTENANCE

## --- BREATH ANALYSIS ---

BLANK TEST	.000	23:09
INTERNAL STANDARD	VERIFIED	23:09
RADIO INTERFERENCE		

Operator Signature

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



DON L CIBERT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/02/2012  
Number 220328  
Expires 10/02/2014

Director of State Public Health Laboratory

Director, Department of Health