



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**

REPORT #6

By Carol Day at 10:43 am, Dec 01, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940234	NAME OF AGENCY Crocker Police Department	DATE OF INSPECTION 11/15/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 108 S Commercial Street, Crocker MO 65452	TIME OF INSPECTION 10:09 pm
---	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED)      DATE AND TIME (from printout) 11/15/2014 22:09

COMPUTER       DETECTOR

PROGRAM       FILTERS

HEATERS SAMPLE CHAMBER 48 °C       QUARTZ STANDARD

FLOW DETECTOR       CALIBRATION

PUMP HIGH SPEED       PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Rep Co Marketing Inc      LOT # 13002      EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C      SIMULATOR SN SD2729      EXP. DATE 09/11/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .098	TEST 2  .097	TEST 3  .097
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Aaron Baker
---------------	--------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230112      06/12/2015	TELEPHONE NUMBER (573) 736-2211
---	------------------------------------

RETURN COMPLETED REPORT TO THE:      Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER:** RepCo Marketing, Inc.

**LOT NUMBER:** 13002

**EXPIRATION DATE:** June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.


This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.


This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**AARON J BAKER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013  
 NUMBER 230112  
 EXPIRES 06/12/2015

*W. W. S.*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paul Vesterberg*  
 Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (FB-10)

MO 580-0771 (6-10)

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
CROCKER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940234  
11/15/14  
22:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 48c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgshijklmno  
pqrstuvwxyz{|}~\*

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
CROCKER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940234  
11/15/14

TESTING OFFICER:  
BAKER/ARRON/J  
OFFICER I.D.: 506  
PERMIT NUMBER: 230112  
EXPIRATION DATE: 06/12/15  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	23:36
INTERNAL STANDARD	VERIFIED	23:36
EXTERNAL STANDARD	.098	23:36
BLANK TEST	.000	23:37
EXTERNAL STANDARD	.097	23:37
BLANK TEST	.000	23:38
EXTERNAL STANDARD	.097	23:38
BLANK TEST	.000	23:39

N = 3  
SIM. = .1  
AVG. = .0973

Operator Signature *Arron J. Baker* 506

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
CROCKER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940234  
11/15/14

ARREST TIME: 23:00  
SUBJECT NAME:  
RFI/TEST  
DOB: 11/11/11 SEX: M  
STATE/D.L.: NA/NA  
ARRESTING OFFICER:  
NA  
OFFICER I.D.: NA  
TESTING OFFICER:  
BAKER/PARON/J  
OFFICER I.D.: 506  
PERMIT NUMBER: 230112  
EXPIRATION DATE: 06/12/15  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	23:52
INTERNAL STANDARD	VERIFIED	23:52
RADIO INTERFERENCE		

Operator Signature \_\_\_\_\_