



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/14/14-CD

REVIEWED REPORT #6
 By Carol Day at 10:08 am, Mar 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940243	NAME OF AGENCY Billing Police	DATE OF INSPECTION 03-06-2014
LOCATION OF INSTRUMENT (STREET AND CITY) Billings Police		TIME OF INSPECTION 2040

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 03-06-2014 @ 2040
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth LOT # 13210 EXP. DATE 07/29/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2274 EXP. DATE 01-15-2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .103	TEST 2 → .102	TEST 3 → .104
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .190
------------	-----------	-------------	-------------	-------------	-----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>Karl Cuschieri</i>	PRINT FULL NAME Karl Cuschieri
TYPE II PERMIT NUMBER/EXPIRATION DATE 220214 / 09/04/2014	TELEPHONE NUMBER 417-744-2582

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET * HARRISBURG, PA 17111-4611 * TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948243
03/05/14
20:40

BAC DATAMASTER SERIAL NUMBER 948243
03/05/14

----- DIAGNOSTIC CHECK -----

TESTING OFFICER:
CUSCHIERI, KARL
OFFICER I.D.: 952
PERMIT NUMBER: 220214
EXPIRATION DATE: 09/24/14
MISCELLANEOUS DATA:

----- SUPERVISOR MODE -----

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

BLANK TEST	.930	20:51
INTERNAL STANDARD	VERIFIED	20:51
EXTERNAL STANDARD	.102	20:51
BLANK TEST	.200	20:52
EXTERNAL STANDARD	.107	20:52
BLANK TEST	.200	20:52
EXTERNAL STANDARD	.104	20:54
BLANK TEST	.000	20:57

N = 3
SIM. = .1
AVG. = .103

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~*

Operator Signature

Karl Cuschieri 952

2208-02

Operator Signature

Karl Cuschieri 952

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940243
03/05/14

ARREST TIME: 00:00
SUBJECT NAME:
RFI/TEST
DOB: 03/06/14 SEX: M
STATE/D.L.: NO/940243
ARRESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
TESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
PERMIT NUMBER: 200214
EXPIRATION DATE: 09/04/14
MISCELLANEOUS DATA:

----- BREATH ANALYSIS -----

BLANK TEST	.000	20:57
INTERNAL STANDARD	VERIFIED	20:57
RADIO INTERFERENCE		

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940240
03/05/14

ARREST TIME: 00:00
SUBJECT NAME:
BLANK/TEST
DOB: 03/06/14 SEX: M
STATE/D.L.: NO/940243
ARRESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 950
TESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
PERMIT NUMBER: 200214
EXPIRATION DATE: 09/04/14
MISCELLANEOUS DATA:

----- BREATH ANALYSIS -----

BLANK TEST	.000	20:59
INTERNAL STANDARD	VERIFIED	20:59
SUBJECT SAMPLE	.000	20:59
BLANK TEST	.000	21:00

Operator Signature

Karl Cuschieri 952

2208-02

Operator Signature

Karl Cuschieri 952

2208-02

State of Missouri
DEPARTMENT OF HEALTH



PERMIT
- TYPE II



KARLA CUSCHIERI

is hereby authorized to instruct and supervise operators, train instructors, install, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, issued under the provisions of sections 577.020 through 577.041, RSMo, 1986.

Date 09/04/2012

Number 220214

Expires 09/04/2014

MO 889-9771 (7-08)

Director of State Public Health Laboratory

Director, Department of Health