MISSOURI DEPARTMENT OF HEA[illegible]ES
STATE PUBLIC HEALTH LABORAT[illegible]
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 2:42 pm, Nov 18, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
940243	Billings Police	11/17/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
Billings Police Department		15:59

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 11/17/2014 @ 15:59
- [x] COMPUTER — [x] DETECTOR
- [x] PROGRAM — [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 49°C — [x] QUARTZ STANDARD
- [x] FLOW DETECTOR — [x] CALIBRATION
- [x] PUMP HIGH SPEED — [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Guth LOT # 14030 EXP. DATE 01/20/2016
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN 093752 EXP. DATE 10/06/2016
- [x] CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.098	.098	.098

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	2	1	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Karl Cuschieri	Karl Cuschieri
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240336 / 09-15-2016	(417) 744-2582

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
690 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5170

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Certificant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940243
11/17/14

TESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
PERMIT NUMBER: 240325
EXPIRATION DATE: 09/15/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	16:02
INTERNAL STANDARD	VERIFIED	16:02
EXTERNAL STANDARD	.098	16:02
BLANK TEST	.000	16:03
EXTERNAL STANDARD	.098	16:03
BLANK TEST	.000	16:04
EXTERNAL STANDARD	.098	16:04
BLANK TEST	.000	16:05

N = 3
SIM. = .1
AVG. = .098

Operator Signature Karl Cuschieri 952

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940243
11/17/14
15:59

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Signature Karl Cuschieri 952

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940243
11/17/14

ARREST TIME: 00:00
SUBJECT NAME:
RFI/TEST
DOB: 11/17/14 SEX: M
STATE/D.L.: MO/940243
ARRESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
TESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
PERMIT NUMBER: 240335
EXPIRATION DATE: 09/15/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	16:07
INTERNAL STANDARD	VERIFIED	16:07
RADIO INTERFERENCE		

Operator Signature Karl Cuschieri 952

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BILLINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940243
11/17/14

ARREST TIME: 00:00
SUBJECT NAME:
BLANK/TEST
DOB: 11/17/14 SEX: M
STATE/D.L.: MO/940243
ARRESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
TESTING OFFICER:
CUSCHIERI/KARL
OFFICER I.D.: 952
PERMIT NUMBER: 240335
EXPIRATION DATE: 09/15/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	16:10
INTERNAL STANDARD	VERIFIED	16:10
SUBJECT SAMPLE	.000	16:11
BLANK TEST	.000	16:11

Operator Signature Karl Cuschieri 952

2208-02

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KARL CUSCHIERI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/15/2014

NUMBER 240335

EXPIRES 9/15/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling ,acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RG-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CUSCHIERI, KARL
Permit No 240335
Date Issued 9/15/2014 **Date Expires** 9/15/2016