



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 12:40 pm, Dec 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 3 months). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940258	NAME OF AGENCY TROY POLICE DEPARTMENT	DATE OF INSPECTION 12/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 800 Cap Au Gris, Troy		TIME OF INSPECTION 1207 Hours

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12/07/14 12:07
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GMH LOT # 13280 EXP. DATE 10/16/15
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2505 EXP. DATE 01/23/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .098%	TEST 2 .099%	TEST 3 .100%
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)


REFUSALS 0	(.0-.04) 2	(.05-.09) 4	(.10-.14) 0	(.15-.19) 0	OVER .19 0
-------------------	-------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME TERRENCE CLAYTON
TYPE II PERMIT NUMBER EXPIRATION DATE 220414 12/27/2014	TELEPHONE NUMBER (636) 528-6179
RETURN COMPLETED REPORT TO THE:	Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901



PERMIT
TYPE II



BERNIE CLAYTON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220414

Expires 12/27/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health

Lab. 4 (R7-BB)



GUTH LABORATORIES, INC.

580 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1217%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 16, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 940258
12/07/14

TESTING OFFICER:
CLAYTON/BERNIE
OFFICER I.D.: 131
PERMIT NUMBER: 220414
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- SUPERVISOR NAME ---

BLANK TEST	.000	12:21
INTERNAL STANDARD	VERIFIED	12:21
EXTERNAL STANDARD	.098	12:22
BLANK TEST	.000	12:22
EXTERNAL STANDARD	.099	12:23
BLANK TEST	.000	12:23
EXTERNAL STANDARD	.100	12:24
BLANK TEST	.000	12:24

N = 3
S.M. = .1
AVG. = .099

Operator Signature

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 940258
12/07/14
12:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefshijklmnop
PQRSTUVWXYZ()++

Operator Signature

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 940258
12/07/14

ARREST TIME: 11:45
SUBJECT NAME:
TEST
JOB: 01/01/01 SEX: M
STATE/D.L.: NO/1234567890
ARRESTING OFFICER:
CLAYTON/BERNIE
OFFICER I.D.: 131
TESTING OFFICER:
CLAYTON/BERNIE
OFFICER I.D.: 131
PERMIT NUMBER: 220414
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 12:31
INTERNAL STANDARD VERIFIED 12:31
SUBJECT SAMPLE .000 12:31
BLANK TEST .000 12:32

Operator Signature

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 940258
12/07/14

ARREST TIME: 11:45
SUBJECT NAME:
RFI/TEST
JOB: 01/01/01 SEX: M
STATE/D.L.: NO/1234567890
ARRESTING OFFICER:
CLAYTON/BERNIE
OFFICER I.D.: 131
TESTING OFFICER:
CLAYTON/BERNIE
OFFICER I.D.: 131
PERMIT NUMBER: 220414
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 12:28
INTERNAL STANDARD VERIFIED 12:28
RADIO INTERFERENCE

Operator Signature