



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 2:44 pm, Dec 03, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 15 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940445	NAME OF AGENCY Holden Police Department	DATE OF INSPECTION 12/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 110 West 3rd Street Holden, Mo.		TIME OF INSPECTION 8:48 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12/01/14 08:48
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Labs. LOT # 14200 EXP. DATE 08/05/2016
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2252 EXP. DATE 02/03/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .100	TEST 3 .101
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Thomas L. Wehr
TYPE II PERMIT NUMBER/EXPIRATION DATE 220435 12/27/2014	TELEPHONE NUMBER (816) 850-4154

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
HOLDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940445
12/01/14
08:48

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (84-87-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	58c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;{<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefg hijklmno
pqrstuvwxyz{|}~

Operator Signature 

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
HOLDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940445
12/01/14


TESTING OFFICER:
MEHR/THOMAS/L
OFFICER I.D.: 006
PERMIT NUMBER: 200405
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	00:50
INTERNAL STANDARD	VERIFIED	00:50
EXTERNAL STANDARD	.099	00:53
BLANK TEST	.000	00:54
EXTERNAL STANDARD	.100	00:54
BLANK TEST	.000	00:55
EXTERNAL STANDARD	.101	00:55
BLANK TEST	.000	00:56

N = 3
SIM. = .1
AVG. = .1

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
HOLDEN POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 940445
12/01/14

ARREST TIME: 09:08
SUBJECT NAME:
R/F/I/TEST
DOB: 12/01/14 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
WEHR/THOMAS/L
OFFICER I.D.: 306
TESTING OFFICER:
WEHR/THOMAS/L
OFFICER I.D.: 306
PERMIT NUMBER: 220405
EXPIRATION DATE: 12/27/14
MISCELLANEOUS DATA:
R/F/I/TEST
R/F/I/TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:59
INTERNAL STANDARD	VERIFIED	08:59
RADIO INTERFERENCE		

Operator Signature



2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



THOMAS L WEHR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220435

Expires 12/27/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health