



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM

RECEIVED
 By Carol Day at 3:00 pm, Jul 28, 2014

DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950001	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/18/2014
LOCATION OF INSTRUMENT (STREET AND CITY) US 169 at Harlem RD, Kansas City		TIME OF INSPECTION 11:34 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/18/14 23:34
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G11068 EXP. DATE 02/14/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← .098%	TEST 2 ← .098%	TEST 3 ← .097%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0	(.0-04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within the guidelines set forth by the Missouri Department of Health.

INSPECTING OFFICER

SIGNATURE Cpl. B. N. Comer, #263	PRINT FULL NAME Cpl. B. N. Comer, #263
TYPE II PERMIT NUMBER/EXPIRATION DATE 240085 03/11/2016	TELEPHONE NUMBER (816) 322-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 950001
07/19/14
20194

--- DIAGNOSTIC CHECK ---

COMPUTERS: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS:
SAMPLE CHAMBER: 49E
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTORS: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~+*

Operator Signature CP. B.M. Comar #263

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 950001
07/19/14

ARREST TIME: 23:10
SUBJECT NAME:
RADIO FREQUENCY/TEST
DOB: 07/07/70 SEX: F
STATE/D.L.I. NO/1234567
ARRESTING OFFICER:
COMER/2/M
OFFICER I.D. # 263
TESTING OFFICER:
COMER/E/M
OFFICER I.D. # 263
PERMIT NUMBER: 240095
EXPIRATION DATE: 03/11/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 23:44
INTERNAL STANDARD VERIFIED 23:44
RADIO INTERFERENCE

Operator Signature CP. B.M. Comar #263

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DRINKMASTER SERIAL NUMBER 950001
07/19/14

TESTING OFFICER:
COMER/B/N
OFFICER I.D.# 263
PERMIT NUMBER: 240095
EXPIRATION DATE: 03/11/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

TEST TYPE	VERIFIED	TIME
BLANK TEST	.000	23:57
INTERNAL STANDARD		23:57
EXTERNAL STANDARD	.098	23:58
BLANK TEST	.000	23:58
EXTERNAL STANDARD	.098	23:59
BLANK TEST	.000	23:59
EXTERNAL STANDARD	.097	00:00
BLANK TEST	.000	00:00

N = 3
SIM. = .1
RMS. = .0976

Operator Signature CP. 181. Com #263