



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 9:33 am, Jun 04, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950086	NAME OF AGENCY Trenton Police Department	DATE OF INSPECTION 06/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 610 Main St Trenton MO		TIME OF INSPECTION 1:15 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 06/04/14 01:18
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo _____ LOT # 13002 _____ EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____ 34 _____ °C SIMULATOR SN _____ SD2739 _____ EXP. DATE 11/26/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➡ .096	TEST 2 ➡ .096	TEST 3 ➡ .096
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	6	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

RePco Solution Lot #13002

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Keith J Edmonston
TYPE II PERMITS NUMBER/EXPIRATION DATE 240264 06/02/2016	TELEPHONE NUMBER (660) 359-5557

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
TRENTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950086
06/04/14
01:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-02-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

#%&'()+,-./0123456789:;<=>@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
PQRSTUVWXYZ{|}~*

Operator Signature



**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
TRENTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950086
06/04/14


TESTING OFFICER:
EDMONSTON/KEITH/J
OFFICER I.D.# 112
PERMIT NUMBER: 240264
EXPIRATION DATE: 06/02/16
MISCELLANEOUS DATA:
TRENTON MAY MAT

--- SUPERVISOR MODE ---

BLANK TEST	.000	01:21
INTERNAL STANDARD	VERIFIED	01:21
EXTERNAL STANDARD	.096	01:21
BLANK TEST	.000	01:22
EXTERNAL STANDARD	.096	01:22
BLANK TEST	.000	01:23
EXTERNAL STANDARD	.096	01:23
BLANK TEST	.000	01:24

N = 3
SIM. = .1
AVG. = .096

Operator Signature



BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
TRENTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950096
05/04/14

ARREST TIME: 01:00
SUBJECT NAME:
DUE-COOKIE/M
DOB: 01/01/41 SEX: M
STATE: MO L.A. NO: 1534567
ARRESTING OFFICER:
EDMONSTON-KEITH/J
OFFICER I.D.: 118
TESTING OFFICER:
EDMONSTON-KEITH/J
OFFICER I.D.: 118
PERMIT NUMBER: 248264
EXPIRATION DATE: 06/02/14
MISCELLANEOUS DATA:
TRENTON MAY RFI

--- BREATH ANALYSIS ---

BLANK TEST	.000	01:27
INTERNAL STANDARD	VERIFIED	01:27
RADIO INTERFERENCE		

Operator Signature





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

KEITH J EDMONSTON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/2/2014

NUMBER 240264

EXPIRES 6/2/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator EDMONSTON, KEITH
Permit No 240264
Date Issued 6/2/2014 Date Expires 6/2/2016