



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

received 1/14/14-cd  
 REPORT #6

**REVIEWED**  
 By Carol Day at 11:41 am, Feb 05, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950115	NAME OF AGENCY FERGUSON POLICE DEPARTMENT	DATE OF INSPECTION 01/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 222 SOUTH FLORISSANT ROAD FERGUSON		TIME OF INSPECTION 11:16 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/07/2014 11:19
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO MARKETING _____ LOT # 13002 _____ EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 _____ °C SIMULATOR SN SD2243 _____ EXP. DATE 11/21/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .098	TEST 2 ☛ .099	TEST 3 ☛ .099
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE <i>P.O. Chris Midyett 502</i>	PRINT FULL NAME P.O. CHRIS MIDYETT 502
TYPE II PERMIT NUMBER/EXPIRATION DATE 230291 11/26/2015	TELEPHONE NUMBER (314) 522-3100

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

# BAC DataMaster

## Evidence Ticket

STATE OF MISSOURI  
BPC DATAMASTER SERIAL NUMBER 950115  
01/07/14


TESTING OFFICER:  
MIDVETT/CHRIS  
OFFICER I.D.: 582  
PERMIT NUMBER: 230291  
EXPIRATION DATE: 11/26/15  
MISCELLANEOUS DATA:  
MONTHLY MAINT.

--- SUPERVISOR MODE ---

TEST TYPE	VERIFIED	TIME
BLANK TEST	.000	11:15
INTERNAL STANDARD	11:15	11:15
EXTERNAL STANDARD	.098	11:15
BLANK TEST	.000	11:16
EXTERNAL STANDARD	.099	11:16
BLANK TEST	.000	11:17
EXTERNAL STANDARD	.099	11:17
BLANK TEST	.000	11:18

N = 3  
SIM. = .1  
RM9. = .0986

Operator Signature



# BAC DataMaster

## Evidence Ticket

STATE OF MISSOURI  
BPC DATAMASTER SERIAL NUMBER 950115  
01/07/14  
11:19

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!@#\$%^&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefg hijklmno  
P-q-rstuvwxyz{|}~\*

Operator Signature



**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
BAC DATAMASTER SERIAL NUMBER 950115  
01/07/14

ARREST TIME: 11:00  
SUBJECT NAME:  
JONES/THOMAS  
DOB: 09/12/44      SEX: M  
STATE/D.L.: MO/987654321  
ARRESTING OFFICER:  
MIDYETT/CHRIS  
OFFICER I.D.: 502  
TESTING OFFICER:  
MIDYETT/CHRIS  
OFFICER I.D.: 502  
PERMIT NUMBER: 230291  
EXPIRATION DATE: 11/26/15  
MISCELLANEOUS DATA:  
RFI

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:23
INTERNAL STANDARD	VERIFIED	11:23
RADIO INTERFERENCE		

Operator Signature

*P. O. J. 502*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**CHRISTOPHER R MIDYETT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/26/2013

NUMBER 230291

EXPIRES 11/26/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator MIDYETT, CHRISTOPHER  
Permit No 230291  
Date Issued 11/26/2013 Date Expires 11/26/2015

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.