



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 11:35 am, Jan 16, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950117	NAME OF AGENCY Brookfield Police Department	DATE OF INSPECTION 01/16/2014
-------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 116 W. Brooks St. Brookfield, MO 64628	TIME OF INSPECTION 3:15 am
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01-16-14 0315</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50°C</u>	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing</u> LOT # <u>12002</u> EXP. DATE <u>08/29/2014</u>

<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>093752</u> EXP. DATE <u>03/29/2014</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➡ <u>.099</u>	TEST 2 ➡ <u>.099</u>	TEST 3 ➡ <u>.099</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 This instrument meets Department of Health regulations.

INSPECTING OFFICER	
SIGNATURE <i>[Signature]</i> 250	PRINT FULL NAME Jonathan G. Bagley
TYPE II PERMIT NUMBER EXPIRATION DATE 240005 01-14-2016	TELEPHONE NUMBER (660) 258-3385

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

REP CO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-8460

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 950117
01/16/14

ARREST TIME: 00:01
SUBJECT NAME: RFI TEST
DOB: 01/01/78 SEX: M
STATE/D.L.: NO-4567891234
ARRESTING OFFICER: BAGLEY/JUN/G

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS: 50c

SAMPLE CHAMBER: OKAY

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

TESTING OFFICER: BAGLEY/JUN/G

OFFICER I.D.: 250

PERMIT NUMBER: 240005

EXPIRATION DATE: 01/14/16

MISCELLANEOUS DATA:

----- SUPERVISOR NODE -----

BLANK TEST	.000	03:22
INTERNAL STANDARD	VERIFIED	03:22
EXTERNAL STANDARD	.099	03:22
BLANK TEST	.000	03:23
EXTERNAL STANDARD	.099	03:23
BLANK TEST	.000	03:24
EXTERNAL STANDARD	.099	03:24
BLANK TEST	.000	03:25

N = 3
SIM. = .1
AVG. = .099

PRINTER TEST
!#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ\^_`abcdefg hijklmno
pqrstuvwxyz{|}~+*

Operator Signature: *[Signature]*

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 950117
01/16/14

TESTING OFFICER: BAGLEY/JUN/G

OFFICER I.D.: 250

PERMIT NUMBER: 240005

EXPIRATION DATE: 01/14/16

MISCELLANEOUS DATA:

----- SUPERVISOR NODE -----

BLANK TEST	.000	03:22
INTERNAL STANDARD	VERIFIED	03:22
EXTERNAL STANDARD	.099	03:22
BLANK TEST	.000	03:23
EXTERNAL STANDARD	.099	03:23
BLANK TEST	.000	03:24
EXTERNAL STANDARD	.099	03:24
BLANK TEST	.000	03:25

N = 3
SIM. = .1
AVG. = .099

Operator Signature: *[Signature]*

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATA MASTER SERIAL NUMBER 950117
01/16/14

ARREST TIME: 00:01

SUBJECT NAME: RFI TEST

DOB: 01/01/78 SEX: M

STATE/D.L.: NO-4567891234

ARRESTING OFFICER: BAGLEY/JUN/G

TESTING OFFICER: BAGLEY/JUN/G

OFFICER I.D.: 250

PERMIT NUMBER: 240005

EXPIRATION DATE: 01/14/16

MISCELLANEOUS DATA:

----- BREATH ANALYSIS -----

BLANK TEST	.000	03:01
INTERNAL STANDARD	VERIFIED	03:02
RADIO INTERFERENCE		

Operator Signature: *[Signature]*



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JONATHAN G BAGLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/14/2014

NUMBER 240005

EXPIRES 1/14/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (05-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BAGLEY, JONATHAN
Permit No 240005
Date Issued 1/14/2014 Date Expires 1/14/2016