MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 1:06 pm, May 07, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
950117	Brookfield Police Department	05/05/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
116 W. Brooks St. Brookfield, MO	10:12 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 05/05/14 10:12
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 50 °C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER RepCo Marketing LOT # 12002 EXP. DATE 08/29/2014
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN 093752 EXP. DATE 03/11/2015
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.096	.096	.096

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	2	0	2	0	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument meets Department of Health standards.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸	Jon Bagley
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240005 01/14/2016	(660) 258-3385

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 12002
EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **12002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1209** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **August 30, 2012** The expiration date for this lot number is **August 29, 2014** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 950117
05/05/14

ARREST TIME: 00:01
SUBJECT NAME:
BAGLEY/JON/G
DOB: 07/13/78 SEX: M
STATE/D.L.: MO/T12345678
ARRESTING OFFICER:
BAGLEY/JON/G
OFFICER I.D.: 250
TESTING OFFICER:
BAGLEY/JON/G
OFFICER I.D.: 250
PERMIT NUMBER: 240005
EXPIRATION DATE: 01/14/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:31
INTERNAL STANDARD	VERIFIED	10:31
RADIO INTERFERENCE		

Operator Signature [signature] 250

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 950117
05/05/14

TESTING OFFICER:
BAGLEY/JON/G
OFFICER I.D.: 250
PERMIT NUMBER: 240005
EXPIRATION DATE: 01/14/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:21
INTERNAL STANDARD	VERIFIED	10:21
EXTERNAL STANDARD	.096	10:21
BLANK TEST	.000	10:22
EXTERNAL STANDARD	.096	10:22
BLANK TEST	.000	10:23
EXTERNAL STANDARD	.096	10:23
BLANK TEST	.000	10:24

N = 3
SIM. = .1
AVG. = .096

Operator Signature [signature] 250

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 950117
05/05/14
10:12

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature [signature] 22

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

JONATHAN G BAGLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/14/2014

NUMBER 240005

EXPIRES 1/14/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BAGLEY, JONATHAN
Permit No 240005
Date Issued 1/14/2014 Date Expires 1/14/2016