



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED  
 By Carol Day at 1:00 pm, Dec 18, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>950117 CD</b>	NAME OF AGENCY Brookfield Police Department	DATE OF INSPECTION 12/14/2014
-----------------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 116 W. Brooks St. Brookfield, MO 64628	TIME OF INSPECTION 4:44 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12/14/14 04:45
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 14001 EXP. DATE 04/30/2016	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN 093752 EXP. DATE 03/11/2015	

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .100	TEST 2 → .101	TEST 3 → .100
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).  
 This simulator and instrument meet Department of Health standards.

INSPECTING OFFICER	
SIGNATURE <i>Jon Bagley</i>	PRINT FULL NAME Jon Bagley
TYPE & PERMIT NUMBER EXPIRATION DATE 240095 01/14/2016	TELEPHONE NUMBER (660) 258-3385

RETURN COMPLETED REPORT TO THE  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 14001**

**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 958117  
12/14/14

TESTING OFFICER:  
BRADLEY/JONAS  
OFFICER I.D.: 250  
PERMIT NUMBER: 240005  
EXPIRATION DATE: 01/14/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 04:49  
INTERNAL STANDARD .000 VERIFIED 04:49  
EXTERNAL STANDARD .100 04:50  
BLANK TEST .000 04:51  
EXTERNAL STANDARD .101 04:51  
BLANK TEST .000 04:52  
EXTERNAL STANDARD .100 04:52  
BLANK TEST .000 04:52

N = 3  
SIM. = .1  
RMS. = .1003

Operator Signature: *Jonas* 250

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 958117  
12/14/14  
04:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 50c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()\*+,-./0123456789:;<=>?@ABDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
PQRSTUVWXYZ ( ) \* +

Operator Signature: *Jonas* 250

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 958117  
12/14/14

ARREST TIME: 00:00  
SUBJECT NAME:  
TEST/RFI  
DOB: 10/16/78 SEX: M  
STATE/D.L.: MO/1746574  
ARRESTING OFFICER:  
BRADLEY/JONAS  
OFFICER I.D.: 250  
TESTING OFFICER:  
BRADLEY/JONAS  
OFFICER I.D.: 250  
PERMIT NUMBER: 240005  
EXPIRATION DATE: 01/14/16  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000  
INTERNAL STANDARD .000 VERIFIED  
RADIO INTERFERENCE

Operator Signature: *Jonas* 250



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**PERMIT  
 TYPE II**

**JONATHAN G BAGLEY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/14/2014

NUMBER 240005

EXPIRES 1/14/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **BAGLEY, JONATHAN**  
 Permit No **240005**  
 Date Issued **1/14/2014**    Date Expires **1/14/2016**