



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 7:28 am, Jan 15, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950122	NAME OF AGENCY Clay County Sheriff's Department	DATE OF INSPECTION 01/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 14 S. Water Street, Liberty, Mo. 64068		TIME OF INSPECTION 11:11 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/14/2014 23:11
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN DR 5389 EXP. DATE 01/16/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .099	TEST 2 • .099	TEST 3 • .101
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	5	(0-.04)	0	(.05-.09)	3	(.10-.14)	8	(.15-.19)	3	OVER .19	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Donnie St. John
---------------	------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230044 03/11/2015	TELEPHONE NUMBER (816) 407-3700
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DONNIE ST JOHN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 03/11/2013

NUMBER 230044

EXPIRES 03/11/2015

MO 680-0771 (6-10)


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


Acting Director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CLAY COUNTY SHERIFF'S DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950122
01/14/14

ARREST TIME: 23:00

SUBJECT NAME:

TEST

DOB: 01/14/14 SEX: M

STATE/D.L.: MO/TEST

ARRESTING OFFICER:

TEST

OFFICER I.D.: TEST

TESTING OFFICER:

STJOHN/DONNIE

OFFICER I.D.: 8129

PERMIT NUMBER: 230044

EXPIRATION DATE: 03/11/15

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	23:20
INTERNAL STANDARD	VERIFIED	23:20
RADIO INTERFERENCE		

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CLAY COUNTY SHERIFF'S DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950122
01/14/14
23:11

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJK
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgghijklmnop
qrstuvwxyz{|}~*

Operator Signature



BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CLAY COUNTY SHERIFF'S DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950122
01/14/14

TESTING OFFICER:
STJOHN/DONNIE
OFFICER I.D.: 8129
PERMIT NUMBER: 230044
EXPIRATION DATE: 03/11/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	23:14
INTERNAL STANDARD	VERIFIED	23:14
EXTERNAL STANDARD	.099	23:14
BLANK TEST	.000	23:15
EXTERNAL STANDARD	.099	23:15
BLANK TEST	.000	23:16
EXTERNAL STANDARD	.101	23:16
BLANK TEST	.000	23:17

N = 3
SIM. = .1 /
AVG. = .0996

Operator Signature 