



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED 3/5/14-CD

**RECEIVED**  
 By Carol Day at 4:03 pm, Mar 20, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950124	NAME OF AGENCY Kirksville Police Department	DATE OF INSPECTION 02/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 119 E McPherson, Kirksville, MO, 63501		TIME OF INSPECTION 11:37 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/28/2014 11:37 am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Labs LOT # 13210 EXP. DATE 07/29/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2271 EXP. DATE 12/19/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .100	TEST 2 • .101	TEST 3 • .102
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 1
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>[Signature]</i> KPD 383	PRINT FULL NAME Juan B Chairez
TYPE 41 PERMIT NUMBER/EXPIRATION DATE 220412 12/27/2014	TELEPHONE NUMBER (660) 785-6945

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

600 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights  
Calibration verification is done prior to each use utilizing NIST traceable weights*

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



JUAN B CHAIREZ

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s).

INTOXILYZER 5000; DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220412

Expires 12/27/2014

MO 549.0771 (1-08)

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health

LSB 419-1261

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

KIRKPATRICK POLICE DEPARTMENT  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION

KIRKPATRICK POLICE DEPARTMENT  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION  
 1000 BRADWOOD DRIVE, SUITE 100  
 KIRKPATRICK, ALABAMA 36020  
 PHONE: 205-336-1111  
 FAX: 205-336-1112  
 WWW: WWW.KIRKPATRICKPOLICE.COM  
 OPERATIONAL SECTION

Operator Signature J.C. KPD 383

Operator Signature J.C. KPD 383

