



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED** REPORT #6  
 By Carol Day at 1:54 pm, Sep 10, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>950125</b>	NAME OF AGENCY <b>MSHP</b>	DATE OF INSPECTION <b>09/07/2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>17201 PARADESIAN, SMITHVILLE, MO</b>		TIME OF INSPECTION <b>07:30</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>09/07/14 07:54</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **REPCO MARKETING INC** LOT # **13001** EXP. DATE **03/07/15**

SIMULATOR TEMP (34°C ± 0.2°C) **34** °C SIMULATOR SN **G6799** EXP. DATE **10/31/14**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • <b>.099</b>	TEST 2 • <b>.101</b>	TEST 3 • <b>.103</b>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS <b>0</b>	(0-.04) <b>0</b>	(.05-.09) <b>0</b>	(.10-.14) <b>0</b>	(.15-.19) <b>0</b>	OVER .19 <b>0</b>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE <i>Cpl COTUBBS</i>	PRINT FULL NAME <b>Corporal Casey O. Tubbs</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240113 03/11/14</b>	TELEPHONE NUMBER <b>816 622 0800</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13001**

**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**CASEY O TUBBS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/11/2014

NUMBER 240113

EXPIRES 3/11/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator TUBBS, CASEY  
 Permit No 240113  
 Date Issued 3/11/2014 Date Expires 3/11/2016

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
MISSOURI STATE WATER PATROL

BAC DATAMASTER SERIAL NUMBER 950125  
09/07/14  
07:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 49c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@##%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefshijklmno  
pqrstuvwxyz{|}~\*\*

OPERATOR SIGNATURE *Glenn Smith*, 1122  
*SMITH, GLENN*

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
MISSOURI STATE WATER PATROL

BAC DATAMASTER SERIAL NUMBER 950125  
09/07/14

TESTING OFFICER:  
TUBBS/C/O

OFFICER I.D.: 1122  
PERMIT NUMBER: 240113  
EXPIRATION DATE: 03/11/16  
MISCELLANEOUS DATA:  
SEPT. MAIN.

--- SUPERVISOR MODE ---

BLANK TEST .000 07:57  
INTERNAL STANDARD VERIFIED 07:57  
EXTERNAL STANDARD .099 07:57  
BLANK TEST .000 07:58  
EXTERNAL STANDARD .101 07:58  
BLANK TEST .000 07:59  
EXTERNAL STANDARD .103 07:59  
BLANK TEST .000 08:00

N = 3  
SIM. = .1  
AVG. = .101

OPERATOR SIGNATURE *Glenn Smith*, 1122  
*SMITH, GLENN*

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
MISSOURI STATE WATER PATROL

BAC DATA MASTER SERIAL NUMBER 950125  
09/07/14

ARREST TIME: 07:00  
SUBJECT NAME:

DOE/JOHN/Z

DOB: 01/01/80

SEX: M

STATE/D.L.#: IL/1234567890

ARRESTING OFFICER:

TUBBS/C/O

OFFICER I.D.#: 1122

TESTING OFFICER:

TUBBS/C/O

OFFICER I.D.#: 1122

PERMIT NUMBER: 240113

EXPIRATION DATE: 03/11/16

MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:04
INTERNAL STANDARD	VERIFIED	
RADIO INTERFERENCE		08:04

OPERATOR SIGNATURE

*Carl Coatsworth, 1122*  
SMITHVILLE