



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 1/28/14-cd

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed in service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 1:16 pm, Feb 04, 2014

DATAMASTER SN 950424	NAME OF AGENCY Slater Police Department	DATE OF INSPECTION 01/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 117 N. Walnut St. Slater, Missouri 65349		TIME OF INSPECTION 2:20 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/14/2014 at 1:51 am
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 34 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13210 EXP. DATE 07/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 3318 EXP. DATE 10/17/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .102	TEST 2 • .102	TEST 3 • .102
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This Datamaster Instrument is operating within the requirements set by the Missouri Department of Health and Senior Services.

INSPECTING OFFICER	
SIGNATURE Joseph E. Valiquette	PRINT FULL NAME Joseph E. Valiquette
TYPE II PERMIT NUMBER/EXPIRATION DATE 230248 10/24/2015	TELEPHONE NUMBER (660) 529-2241

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JOSEPH E VALIQUETTE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/19/2012

Number 220371

Expires 10/19/2014

Director of State Public Health Laboratory

Director, Department of Health

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SLATER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950424
01/14/14

TESTING OFFICER:
VALIQUETTE/JOSEPH/E
OFFICER I.D.: 242
PERMIT NUMBER: 230248
EXPIRATION DATE: 10/24/15
MISCELLANEOUS DATA:
N/A
N/A

--- SUPERVISOR MODE ---

BLANK TEST	.000	02:04
INTERNAL STANDARD	VERIFIED	02:04
EXTERNAL STANDARD	.102	02:04
BLANK TEST	.000	02:05
EXTERNAL STANDARD	.102	02:05
BLANK TEST	.000	02:06
EXTERNAL STANDARD	.102	02:06
BLANK TEST	.000	02:07

N = 3
SIM. = .1
AVG. = .102

OPERATOR SIGNATURE

Stock No.

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SLATER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950424
01/14/14

ARREST TIME: 01:50
SUBJECT NAME:
VALIQUETTE/JOSEPH/E
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
VALIQUETTE/JOSEPH/E
OFFICER I.D.: 242
TESTING OFFICER:
VALIQUETTE/JOSEPH/E
OFFICER I.D.: 242
PERMIT NUMBER: 230248
EXPIRATION DATE: 10/24/15
MISCELLANEOUS DATA:
RFI TEST
N/A

--- BREATH ANALYSIS ---

BLANK TEST	.000	02:10
INTERNAL STANDARD	VERIFIED	02:10
RADIO INTERFERENCE		

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SLATER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 950424
01/14/14
01:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefg hijklmno
pqrstuvwxyz{|}~

OPERATOR SIGNATURE

Joseph E. Uslijette

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901