MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 10:41 am, Jun 11, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
990040	Republic Police Department	05/29/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
540 Civic Blvd Republic	2:45 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 05/29/2014 @ 3:11 PM
- ☑ COMPUTER — ☑ DETECTOR
- ☑ PROGRAM — ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 49 °C — ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR — ☑ CALIBRATION
- ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories — LOT # 13280 — EXP. DATE 10/16/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C — SIMULATOR SN SD3326 — EXP. DATE 08/07/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.097	.097	.097

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	1	0	0	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Jennifer L. Stephens #097	Jennifer L. Stephens
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
230170 08/14/15	(417) 732-3979

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

PERMIT

TYPE II

JENNIFER L STEPHENS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s)

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo

DATE 8/14/2013

NUMBER 230170

EXPIRES 8/14/2015

D 580-0771 (E 10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

(AE-4 (RE 10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator STEPHENS, JENNIFER
Permit No 230170
Date Issued 8/14/2013 Date Expires 8/14/2015

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13280** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1217%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 16, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
REPUBLIC POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 990040
05/29/14

TESTING OFFICER:
STEPHENS/JENNIFER/L
OFFICER I.D.: 097
PERMIT NUMBER: 230170
EXPIRATION DATE: 08/14/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	15:14
INTERNAL STANDARD	VERIFIED	15:14
EXTERNAL STANDARD	.097	15:14
BLANK TEST	.000	15:15
EXTERNAL STANDARD	.097	15:15
BLANK TEST	.000	15:16
EXTERNAL STANDARD	.097	15:16
BLANK TEST	.000	15:17

N = 3
SIM. = .1
AVG. = .097

Operator Signature J. L. Stephens #097

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
REPUBLIC POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 990040
05/29/14
15:11

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```

Operator Signature J. L. Stephens #097

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
REPUBLIC POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 990040
05/29/14

ARREST TIME: 15:00
SUBJECT NAME:
MAY/TEST
DOB: 09/18/78 SEX: M
STATE/D.L.: MO/102938475G
ARRESTING OFFICER:
STEPHENS/JENNIFER/L
OFFICER I.D.: 097
TESTING OFFICER:
STEPHENS/JENNIFER/L
OFFICER I.D.: 097
PERMIT NUMBER: 230178
EXPIRATION DATE: 08/14/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

2208-02

IN THE CIRCUIT COURT OF GREENE COUNTY, MISSOURI
MUNICIPAL DIVISION AT REPUBLIC

AFFIDAVIT

STATE OF MISSOURI)
) SS
COUNTY OF GREENE)

Before me, the undersigned authority, personally appeared __Jennifer Stephens___, who, being by me duly sworn, deposed as follows:

My name is__ Jennifer Stephens___, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of the records of the Republic Police Department Datamaster. Attached hereto is/are 5 page(s) consisting of the following records that are kept by the Republic Police Department in the regular course of business, and it was the regular course of business of the Republic Police Department for an employee or representative of the Republic Police Department with knowledge of the act, event, condition, opinion, or diagnosis recorded to make the record or to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion or diagnosis. The records attached are the original or exact duplicates of the originals of the 05/29/2014 Datamaster Report.

Jennifer L. Stephens #097
Custodian of Records

In witness whereof I have hereunto subscribed my name and affixed my official seal this 29th day of May . 2014 .

Jennifer L Dishman
Notary Public

My Commission expires 7-30-2017

JENNIFER L. DISHMAN
Notary Public - Notary Seal
State of Missouri
Commissioned for Greene County
My Commission Expires: July 30, 2017
Commission Number: 13422794