



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 11:42 am, Apr 03, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 990044	NAME OF AGENCY Cape Girrdeau P.D.	DATE OF INSPECTION 04/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 40 S. Sprigg Cape Girardeau (Interview room)		TIME OF INSPECTION 9:14 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/01/2014 09:14
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13802 EXP. DATE 08/12/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2221 EXP. DATE 07/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .081	TEST 2 → .082	TEST 3 → .082
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>Kevin L. Eudy #106</i>	PRINT FULL NAME Kevin L. Eudy
TYPE II PERMIT NUMBER/EXPIRATION DATE 220132 06/12/2014	TELEPHONE NUMBER (573) 335-6621

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

REPco MARKETING INC.

510-182 STONYERODK DRIVE
RALEIGH, N.C. 27604
919-576-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13802
EXPIRATION DATE: August 12, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13802 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .0968 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .080 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 13, 2013
The expiration date for this lot number is August 12, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

STATE OF ILLINOIS
DEPARTMENT OF HEALTH



PERMIT
TYPE II



KEYWORD STUDY

~~It is hereby authorized to conduct and operate operations, with equipment, located
within the State of Illinois, and within the following listed boundaries:~~

DATABACKER

~~in the State of Illinois, in the County of Cook, State of Illinois, and within the following listed boundaries:
as shown under the jurisdiction of sections 577.020 through 577.040, Public Act 1985~~

05/22/2012

220132

05/22/2014

~~1-6-1985~~

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SENIO LAW ENFORCEMENT ACADEMY

BAC DATAMASTER SERIAL NUMBER 990044
04/01/14

TESTING OFFICER:
EUDY/KEVIN/L
OFFICER I.D.: 106
PERMIT NUMBER: 220132
EXPIRATION DATE: 06/12/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:17
INTERNAL STANDARD	VERIFIED	09:17
EXTERNAL STANDARD	.001	09:17
BLANK TEST	.000	09:18
EXTERNAL STANDARD	.002	09:18
BLANK TEST	.000	09:19
EXTERNAL STANDARD	.002	09:19
BLANK TEST	.000	09:20

N = 3
SIN. = .098
AVG. = .0810

Kevin L Eudy

Signature _____

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SENIO LAW ENFORCEMENT ACADEMY

BAC DATAMASTER SERIAL NUMBER 990044
04/01/14

ARREST TIME: 09:00
SUBJECT NAME:
RFI/TEST
JOB: 12/23/40 RES: F
STATE/D.L.: MO/985767
ARRESTING OFFICER:
N/A

OFFICER I.D.: 1
TESTING OFFICER:
EUDY/KEVIN/L
OFFICER I.D.: 106
PERMIT NUMBER: 220132
EXPIRATION DATE: 06/12/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:20
INTERNAL STANDARD	VERIFIED	09:20
RADIO INTERFERENCE		

Kevin L Eudy

Operator Signature _____

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SAND LAW ENFORCEMENT (MO) (MO) (MO)

BAC DATAMASTER SERIAL NUMBER 5980044
04/01/14
09:14

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS:	
SAMPLE CHAMBER:	48L
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#\$%^&*() -+0123456789:;<=>?@ABCD EFGHIJKL
MNOPQRSTUVWXYZ [] ^ _ ` abc def ghi jkl mno
pqrstuvwxyz { | } ~

Operator Signature _____

11/11/14 #106