



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 10:11 am, Apr 11, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 990044	NAME OF AGENCY Cape Girardeau P.D.	DATE OF INSPECTION 04/08/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 40 S. Sprigg Cape Girardeau (Interview room)		TIME OF INSPECTION 9:14 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/08/2014 09:14
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repco. Marketing Inc. LOT # 13802 EXP. DATE 08/12/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2221 EXP. DATE 07/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➤ .081	TEST 2 ➤ .084	TEST 3 ➤ .084
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE ▶	PRINT FULL NAME Kevin L. Eudy
TYPE II PERMIT NUMBER/EXPIRATION DATE 220132 06/12/2014	TELEPHONE NUMBER (573) 335-6621

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

REP CO MARKETING INC.

3101-188 STONEYBROOK DRIVE
RALEIGH, N.C. 27604
919-576-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13802

EXPIRATION DATE: August 12, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13802 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .0968 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .080 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 13, 2013.
The expiration date for this lot number is August 12, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

STATE OF MASSACHUSETTS
DEPARTMENT OF HEALTH



PERMIT
TYPE II



KEVIN LEIDY

is hereby authorized to install and operate equipment, test methods, assays, methods, procedures and reports, and operate the following branch laboratory:

DATA MASTER

for the determination of the alcohol content of blood from a sample of stored (specimen) at levels under the provisions of sections 577.020 through 577.041, Public Law

05/12/2012

220132

05/12/2014

EX-104149

EX-104149

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SEND LAW ENFORCEMENT AGENCY

BAC DATAMASTER SERIAL NUMBER 000044
04/00/14

ARREST TIME: 09:00
SUBJECT NAME:
RFI/TEST
DOB: 05/23/56 SEX: F
STATE/D.L.# NO/59403737
ARRESTING OFFICER:
N/A
OFFICER I.D.# 1
TESTING OFFICER:
EUDY/RE/TH/L
OFFICER I.D.# 100
PERMIT NUMBER: 003107
EXPIRATION DATE: 06-12-14
MISCELLANEOUS DATA:

DRUG ANALYSIS

BLANK TEST	LOGG	03:24
INTERNAL STANDARD	VERIFIED	03:24
RADIO INTERFERENCE		

Operator Signature *[Signature]*

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SEND LAW ENFORCEMENT AGENCY

BAC DATAMASTER SERIAL NUMBER 000044
04/00/14
09:14

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAMS (04 07 2800):	OKAY
HEATERS:	
SAMPLE CARRIER:	OK
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STABILIZED:	OKAY
CALIBRATION:	OKAY

PRINTED TEST

1980121044: 1403403737: 1.0000000000
ABCDEFGHIJKLMN OPQRSTU VWXYZ0123456789: abcdefghijklmnop
qrstuvwxyz0123456789

Operator Signature *[Signature]*

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
JEMO LAW ENFORCEMENT ACADEMY

BAC DATA MASTER SERIAL NUMBER 090014
04/00/14

TESTING OFFICER:
EUDY KEVIN L
OFFICER I.D. # 166
PERMIT NUMBER: 200170
EXPIRATION DATE: 06/10/14
MISCELLANEOUS DATA:

SUPERVISOR WOOD

BLANK TEST	.000	09:17
INTERNAL STANDARD	VERIFIED	09:17
EXTERNAL STANDARD	.081	09:18
BLANK TEST	.000	09:18
EXTERNAL STANDARD	.004	09:19
BLANK TEST	.000	09:19
EXTERNAL STANDARD	.004	09:20
BLANK TEST	.000	09:20

N = 3
SIM. = .090
AVG. = .083

Operator Signature _____

[Handwritten Signature] 4/10/14