



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 1:04 pm, Jul 21, 2014  
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 990044	NAME OF AGENCY Cape Girardeau Police Dept.	DATE OF INSPECTION 07/09/2014
LOCATION OF INSTRUMENT (STREET AND CITY) BAT Van Cape Girardeau MO		TIME OF INSPECTION 1:18 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/09/14 13:18
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc LOT # 13802 EXP. DATE 08/12/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2221 EXP. DATE 07/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .083	TEST 2 .081	TEST 3 .079
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>alkec #106</i>	PRINT FULL NAME Cpl. Kevin L. Eudy DSN 106
TYPE II PERMIT NUMBER/EXPIRATION DATE 240290 06/23/2016	TELEPHONE NUMBER (573) 335-6621

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

REP CO MARKETING INC.

3101-185 STONEYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-576-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.  
LOT NUMBER: 13802  
EXPIRATION DATE: August 12, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13802 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .0968 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .080 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 13, 2013.  
The expiration date for this lot number is August 12, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**KEVIN L EUDY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/23/2014

NUMBER 240290

EXPIRES 6/23/2016

MO 589-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator EUDY, KEVIN  
Permit No 240290  
Date Issued 6/23/2014 Date Expires 6/23/2016

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
SEMO LAW ENFORCEMENT ACADEMY

BAC DATAMASTER SERIAL NUMBER 990044  
07/09/14

ARREST TIME: 13:00  
SUBJECT NAME:  
RFI/TEST  
DOB: 04/21/45 SEX: M  
STATE/D.L.: MO/40055877  
ARRESTING OFFICER:  
N/A  
OFFICER I.D.: 1  
TESTING OFFICER:  
EUDY/KEVIN/L  
OFFICER I.D.# 106  
PERMIT NUMBER: 240290  
EXPIRATION DATE: 06/23/16  
MISCELLANEOUS DATA:

### BREATH ANALYSIS

BLANK TEST	.000	13:37
INTERNAL STANDARD	VERIFIED	13:37
RADIO INTERFERENCE		

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
SEMO LAW ENFORCEMENT ACADEMY

BAC DATAMASTER SERIAL NUMBER 990044  
07/09/14  
13:18

### DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM (04-87-2809):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;{=}~?@ABCDEFGHI  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~>\*

Operator Signature WKE 2706

Printed on recycled paper with agri-based inks

CMSU 2208-96

Operator Signature WKE 2706

Printed on recycled paper with agri-based inks

CMSU 2208-96

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
SERVO LAW ENFORCEMENT ACADEMY

BAC DATAMASTER SERIAL NUMBER 990044  
87/89/14

TESTING OFFICER:  
EUNY/KEVIN/L  
OFFICER I.D.: 106  
PERMIT NUMBER: 240298  
EXPIRATION DATE: 06/23/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	13:30
INTERNAL STANDARD	VERIFIED	13:30
EXTERNAL STANDARD	.000	13:30
BLANK TEST	.000	13:31
EXTERNAL STANDARD	.001	13:31
BLANK TEST	.000	13:32
EXTERNAL STANDARD	.079	13:32
BLANK TEST	.000	13:30

N = 3  
SIM. = .008  
AVG. = .001

Operator Signature *Elke C #15*

Printed on recycled paper with agri-based inks

CMSU 2208-96