MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED 3/14/14-CD

REPORT #6

REVIEWED
By Carol Day at 2:55 pm, Mar 31, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
990218	Fayette Police Dept.	03/04/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
100 N. Mulberry Fayette Mo 65248	2:31 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 03/04/14 14:31
- [x] COMPUTER — [x] DETECTOR
- [x] PROGRAM — [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 49 °C — [x] QUARTZ STANDARD
- [x] FLOW DETECTOR — [x] CALIBRATION
- [x] PUMP HIGH SPEED — [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc LOT # 13001 EXP. DATE 03/07/15
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD1115 EXP. DATE 01/14/2015
- [x] CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.099	.099	.100

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	2	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operating within Dept. of Health Regulations.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Greg T Lanham	Greg T. Lanham

TYPE II PERMIT NO.	ISSUE/EXPIRATION DATE	TELEPHONE NUMBER
230010	01/21/2015	(660) 248-2241

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1215** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **March 8, 2013**
The expiration date for this lot number is **March 7, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

GREG T LANHAM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/21/2013

NUMBER 230010

EXPIRES 01/21/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 990218
03/04/14

TESTING OFFICER:
LANHAM/GREG/T
OFFICER I.D.: 2
PERMIT NUMBER: 230010
EXPIRATION DATE: 01/21/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	14:38
INTERNAL STANDARD	VERIFIED	14:38
EXTERNAL STANDARD	.099	14:39
BLANK TEST	.000	14:39
EXTERNAL STANDARD	.099	14:40
BLANK TEST	.000	14:41
EXTERNAL STANDARD	.100	14:41
BLANK TEST	.000	14:42

N = 3
SIM. = .1
AVG. = .0993

Signature Greg T Lanham #2

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 990218
03/04/14

ARREST TIME: 14:15
SUBJECT NAME:
SAMPLE
DOB: 09/03/90 SEX: M
STATE D.L.: MO/000000000
ARRESTING OFFICER:
LANHAM/GREG/T
OFFICER I.D.: 2
TESTING OFFICER:
LANHAM/GREG/T
OFFICER I.D.: 2
PERMIT NUMBER: 230010
EXPIRATION DATE: 01/21/15
MISCELLANEOUS DATA:

BREATH ANALYSIS

BLANK TEST	.000	14:3[illegible]
INTERNAL STANDARD	VERIFIED	14:3[illegible]
RADIO INTERFERENCE		

Operator Signature Greg T Lanham #2

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
FAYETTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER [illegible]
03/04/14
14:3[illegible]

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM (04 07 2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	4[illegible]
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
[illegible]

Operator Signature [illegible]