



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 12:33 pm, Oct 07, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 990218	NAME OF AGENCY Fayette Police Dept.	DATE OF INSPECTION 10/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 100 N. Mulberry, Fayette Mo 65248		TIME OF INSPECTION 12:33 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 10/01/2014 12:33
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Repcos Marketing Inc.</u> LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34</u> °C SIMULATOR SN <u>SD1115</u> EXP. DATE <u>01/14/2015</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← .098	TEST 2 ← .098	TEST 3 ← .099
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	3	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operating with in Dept. of Health regulations.

INSPECTING OFFICER	
SIGNATURE <i>Greg T Lanham</i>	PRINT FULL NAME Greg T. Lanham
TYPE II PERMIT NUMBER/EXPIRATION DATE 230010 01/21/2015	TELEPHONE NUMBER (660) 248-2241

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

GREG T LANHAM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/21/2013

NUMBER 230010

EXPIRES 01/21/2015

MO 890-0/71 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (98-10)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
EVIDENCE POLICE DEPARTMENT

DEF. DATA MASTER SERIAL NUMBER 200010
10-01-14

ARREST TIME: 10:17
SUBJECT NAME:
SAMPLE
DOB: 07-04-1974 GENE R
STATE/ALLI. NO: 0000000000
ARRESTING OFFICER:
LAWMAN TREGENT
OFFICER I.D.# 2
TESTING OFFICER:
LAWMAN TREGENT
OFFICER I.D.# 2
PERMIT NUMBER: 200010
EXPIRATION DATE: 01-01-15
MISCELLANEOUS DATA:

BREATH ANALYSIS

BLANK TEST	.000	10:37
INTERNAL STANDARD	VERIFIED	10:37
RADIO INTERFERENCE		

Operator Signature

Greg T. Laska #2

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
EVIDENCE POLICE DEPARTMENT

DEF. DATA MASTER SERIAL NUMBER 200010
10-01-14

TESTING OFFICER:
LAWMAN TREGENT
OFFICER I.D.# 2
PERMIT NUMBER: 200010
EXPIRATION DATE: 01-01-15
MISCELLANEOUS DATA:

CHYMOLOGICAL INFO

BLANK TEST	.000	10:40
INTERNAL STANDARD	VERIFIED	10:40
EXTERNAL STANDARD	.000	10:41
BLANK TEST	.000	10:42
EXTERNAL STANDARD	.000	10:43
BLANK TEST	.000	10:43
INTERNAL STANDARD	.000	10:43
BLANK TEST	.000	10:44

TIME: 10:44
DATE: 10-01-14

Operator Signature

Greg T. Laska #2

BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI
POLICE DEPARTMENT

APP. DATAMASTER SERIAL NUMBER: 1500119
1979 11 14
12498

DIAGNOSTIC CHECK

- COMPUTER: OK
- PROGRAM (94 AT 2000): OK
- VENTS: OK
- SAMPLE CHAMBER: OK
- FLOW DETECTOR: OK
- PUMP: OK
- HIGH SPEED: OK
- DETECTORS: OK
- TEST TUBE: OK
- QUALITY STANDARD: OK
- 70 TREATMENT: OK

PRINTED TEST

DATE: 11/14/79 12:49
TIME: 12:49
TEST NUMBER: 1500119

Operator Signature Greg T. Lanham #2