



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
By Carol Day at 5:48 am, Mar 02, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>030794</b>	PRINTER SN <b>84.9324.031</b>	DATE OF INSPECTION <b>2/27/2015</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>1781 Zumbeth Rd., St. Charles, MO 63303</b>		TIME OF INSPECTION <b>1609hrs</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) **ok**
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **ok**
- PRINTER WORKING PROPERLY **ok**
- TIME AND DATE DISPLAYING PROPERLY **ok**

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER **Guth Laboratories** LOT # **14030** EXP. DATE **1/20/2016**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD2248** SIMULATOR EXP DATE **1/15/2016**
- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <b>.102</b>	TEST 2 • <b>.102</b>	TEST 3 • <b>.101</b>
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING **ok**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>2</b>	(.0-.04) <b>0</b>	(.05-.09) <b>1</b>	(.10-.14) <b>2</b>	(.15-.19) <b>4</b>	(OVER .19) <b>1</b>
-------------------	-------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <i>Daniel Gibbons</i> 324	PRINT NAME <b>Daniel Gibbons 324</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240034 2/11/2016</b>	TELEPHONE NUMBER <b>(636) 949-3300 x2324</b>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of  $0.100 \text{ g}/210\text{L} \pm 3\%$ .

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 01033

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
02/27/15 16:09 .000  
Calibration Check:  
23 02/27/15 16:09 .102

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbach Rd.

St Charles, Mo 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 01034

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
02/27/15 16:11 .000  
Calibration Check:  
24 02/27/15 16:11 .102

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbach Rd.

St. Charles, Mo 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 01035

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
02/27/15 16:13 .000  
Calibration Check:  
25 02/27/15 16:13 .101

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbach Rd.

St Charles, Mo 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 01036

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
12 02/27/15 16:14

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbach Rd.

St Charles, Mo 63303



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**DANIEL P GIBBONS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/20/2015

NUMBER 250021

EXPIRES 1/20/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator GIBBONS, DANIEL  
Permit No 250021  
Date Issued 1/20/2015 Date Expires 1/20/2017