



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance. Send copy to Department of Health and Senior Services; retain original in department file. **By Carol Day at 7:36 am, May 12, 2015**

ALCO SENSOR IV SN 087963	PRINTER SN 096,3580,943	DATE OF INSPECTION 5/6/15
LOCATION OF INSTRUMENT (STREET AND CITY) 2101 Jeffco Blvd Arnold		TIME OF INSPECTION 0348

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **26°/25°/25°**
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER Guth Labs	LOT # 14220 EXP. DATE 9/24/16
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0	SIMULATOR SN 5B1909 SIMULATOR EXP DATE 2/9/15

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 - 1100	TEST 2 - 1102	TEST 3 - 1101
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 1	(.10-.14) 0	(.15-.19) 2	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE [Signature]	PRINT NAME P.O. Jason Ober
TYPE II PERMIT NUMBER/EXPIRATION DATE 230305 12/11/15	TELEPHONE NUMBER (636) 296-2222

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON W O'BARR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2013

NUMBER 230305

EXPIRES 12/11/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator O'BARR, JASON
Permit No 230305
Date Issued 12/11/2013 Date Expires 12/11/2015

PS 70 Serial no. 882963
Version no. 1370

TEST RECORD 08100

Time Date Time 2101

air intake
25 03200/25 03200 .800

Exhaust gas (14.7%)
25 03200/25 03200 .151

Subject Name

Test

Subject I. H.

Test

Operator Name: J. H.

PO J O'Bar 128

Location

2101 Jaffco Blvd

PS 70 Serial no. 882963
Version no. 1370

TEST RECORD 08100

Time Date Time 2101

air intake
25 03200/25 03200 .800

Exhaust gas (14.7%)
25 03200/25 03200 .152

Subject Name

Test

Subject I. H.

Test

Operator Name: J. H.

PO J O'Bar 128

Location

2101 Jaffco Blvd

PS 70 Serial no. 882963
Version no. 1370

TEST RECORD 08100

Time Date Time 2101

air intake
25 03200/25 03200 .800

Subject Name

Test

Subject I. H.

Test

Operator Name: J. H.

PO J O'Bar 128

Location

2101 Jaffco Blvd

PS 70 Serial no. 882963
Version no. 1370

TEST RECORD 08100

Time Date Time 2101

air intake
25 03200/25 03200 .800

Exhaust gas (14.7%)
25 03200/25 03200 .153

Subject Name

Test

Subject I. H.

Test

Operator Name: J. H.

PO J O'Bar 128

Location

2101 Jaffco Blvd

State of Missouri)
)
COUNTY OF JEFFERSON) SS.

AFFIDAVIT

Before me, the undersigned authority, personally appeared, **Patn. Jason O'Barr**, who being by me duly sworn, deposed as follows.

My name is **Patn. Jason O'Barr #128**, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of the records of the **BREATHALYZER MAINTENANCE RECORDS FOR THE ARNOLD POLICE DEPARTMENT**. Attached hereto are 5 pages of records from the **ARNOLD POLICE DEPARTMENT**. These 5 pages of records are kept by the **ARNOLD POLICE DEPARTMENT** in the course of business, and it was the regular course of business of the **ARNOLD POLICE DEPARTMENT** for an employee or representative of the **ARNOLD POLICE DEPARTMENT**, with the knowledge of the act, event, condition, opinion, or diagnosis recorded to make the record to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion, or diagnosis. The records attached hereto are the original or exact duplicates of the original.

Patn. Jason O'Barr
Affiant

In witness whereof I have hereunto subscribed my name and affixed my official seal this 7th day of May, 2015.

Ruth H. Robinson
(Signed)

(Seal)

RUTH H. ROBINSON
Notary Public - Notary Seal
State of Missouri
Commissioned for Jefferson County
My Commission Expires: June 17, 2017
Commission Number: 13450067