



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE

**RECEIVED**  
By Carol Day at 9:02 am, Dec 28, 2015

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087970	PRINTER SN 08C.3527.187	DATE OF INSPECTION 12/16/2015
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 9701 Marion Park Drive, Kansas City, MO 64137	TIME OF INSPECTION 7:19 pm
---	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG521003 EXP. DATE 07/29/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  0.101	TEST 2  0.101	TEST 3  0.101
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Kori Smeiska
---------------	----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 250133/6-8-2017	TELEPHONE NUMBER (816) 234-5000
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

KORI SMEISKA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250133

EXPIRES 6/8/2017

MO 580-0771 (6-10)

*W. V. Vesterly*  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Gail Vesterly*  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-1 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator SMEISKA, KORI  
Permit No 250133  
Date Issued 6/8/2015 Date Expires 6/8/2017

# Airgas

Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

Certificate of Analysis

Customer Name  
*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 3-Aug-2015

Lot # AG521003 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
29-Jul-2017	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010681	391.8 ppm	EB0010603	392.5 ppm
EB0010670	259.8 ppm	EB0010569	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
Date: 2015.08.04 09:55:31 -0500  
Reason: Dry gas standard certification of analysts  
Location: Airgas USA LLC (Lab)

Analyst: *Rod Marsala*  
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 087970  
Version no: 532B

TEST RECORD 00075

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/16/15 19:19 .000  
Calibration Check:  
24 12/16/15 19:19 .101

Subject Name

Subject I.D.

Operator Name, I.D.

*P.O. Smeiska #5260*

Location

*DUI Section*

..

AS IV Serial no: 087970  
Version no: 532B

TEST RECORD 00076

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/16/15 19:21 .000  
Calibration Check:  
25 12/16/15 19:21 .101

Subject Name

Subject I.D.

Operator Name, I.D.

*P.O. Smeiska #5260*

Location

*DUI Section*

..

AS IV Serial no: 087970  
Version no: 532B

TEST RECORD 00077

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/16/15 19:23 .000  
Calibration Check:  
26 12/16/15 19:23 .101

Subject Name

Subject I.D.

Operator Name, I.D.

*P.O. Smeiska #5260*

Location

*DUI Section*

..

AS IV Serial no: 087970  
Version no: 532B

TEST RECORD 00078

Temp Date Time <sup>s/</sup> 210L

VOID: REI  
12 12/16/15 19:24

Subject Name

Subject I.D.

Operator Name, I.D.

*P.O. Smeiska #5260*

Location

*DUI Section*