



**RECEIVED**

By Ellen Strawsine at 4:01 pm, Dec 22, 2015

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN Nixa 095961	PRINTER SN Nixa 099.3586.643	DATE OF INSPECTION 12/14/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 715 W. Center Circle, Nixa		TIME OF INSPECTION 9:30 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing Inc LOT # 14001 EXP. DATE 04/30/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2731 SIMULATOR EXP DATE 08/12/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100

TEST 2 .101

TEST 3 .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	1	(.10-.14)	1	(.15-.19)	2	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT MEETS OR EXCEEDS DOH RULES AND REGULATIONS

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME DARREN WHISNANT
TYPE II PERMIT NUMBER/EXPIRATION DATE 250195 08/18/2017	TELEPHONE NUMBER (417) 725-2510

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

# Nixa Police Department

## Blank (Zero) test Evidence slip

AS IV Serial no: 095961  
Version no: 332B

TEST RECORD 00094

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/14/15 09:54 .000  
Subject Test: Auto  
21 12/14/15 09:54 .000

Subject Name

Subject I.D.

*D. Williams* 421

Operator Name, I.D.

*NIXA P.D.*

Location

!

# Nixa Police Department

## Calibration Check slip's

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00892

Temp	Date	Time	g/	210L
------	------	------	----	------

Air Blank:	12/14/15	09:38	.000	
Calibration Check:	22	12/14/15	09:38	.100

Subject Name

Subject I.D.

*Whinnant* 421  
Operator Name, I.D.

Nixa P.D.  
Location

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00891

Temp	Date	Time	g/	210L
------	------	------	----	------

Air Blank:	12/14/15	09:34	.000	
Calibration Check:	21	12/14/15	09:34	.101

Subject Name

Subject I.D.

*Whinnant* 421  
Operator Name, I.D.

Nixa P.D.  
Location

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00890

Temp	Date	Time	g/	210L
------	------	------	----	------

Air Blank:	12/14/15	09:33	.000	
Calibration Check:	21	12/14/15	09:33	.100

Subject Name

Subject I.D.

*Whinnant* 421  
Operator Name, I.D.

Nixa P.D.  
Location

# Nixa Police Department

## RFI Evidence slip

AS IV Serial no: 095861  
Version no: 532B

TEST RECORD 00893

Temp Date Time 210L

VOID: RFI  
12 12/14/15 09:39

Subject Name

Subject I.D.

*D. Whiteman* 421

Operator Name, I.D.

*Nixa P.D.*

Location



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DARREN WHISNANT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/18/2015

NUMBER 250195

EXPIRES 8/18/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator WHISNANT, DARREN  
 Permit No 250195  
 Date Issued 8/18/2015 Date Expires 8/18/2017

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 14001**

**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.