MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
By Carol Day at 12:22 pm, Aug 21, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in d[illegible]

ALCO SENSOR IV SN: 097403	PRINTER SN: 099.3586.830	DATE OF INSPECTION: 08/18/2015
LOCATION OF INSTRUMENT (STREET AND CITY): 1414 Evergreen, Wellston, MO 63133		TIME OF INSPECTION: 5:34 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Laboratories LOT # 14220 EXP. DATE 09/24/2016

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN 093752 SIMULATOR EXP DATE 04/01/2016

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .099	TEST 3 ☛ .098

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	2	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE: Sgt J Martin DSN: 220	PRINT NAME: Sgt. J. Martin DSN: 220
TYPE II PERMIT NUMBER/EXPIRATION DATE: 250089/ 05/11/2017	TELEPHONE NUMBER: (314) 428-7373

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
Services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
690 NORTH 67th STREET • HARRISBURG, PA 17111-4111 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N0030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN0805130I whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00083

Temp	Date	Time	210L

VOID: RFI
12 08/18/15 05:40

Subject Name
JON DOE
Subject I.D.
123 DRIVE
Operator Name, I.D.
1414 EVERGREEN
Location

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00082

Temp	Date	Time	210L

Air Blank:
08/18/15 05:38 .000
Subject Test: Auto
26 08/18/15 05:38 .098

Subject Name
JON DOE
Subject I.D.
123 DRIVE
Operator Name, I.D.
1414 EVERGREEN
Location

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00081

Temp	Date	Time	210L

Air Blank:
08/18/15 05:36 .000
Subject Test: Auto
25 08/18/15 05:36 .099

Subject Name
JON DOE
Subject I.D.
123 DRIVE
Operator Name, I.D.
1414 EVERGREEN
Location

TEST RECORD - REPRINT

TEST RECORD 00080

Temp	Date	Time	210L

Air Blank:
08/18/15 05:34 .000
Subject Test: Auto
25 08/18/15 05:34 .100

Subject Name
JON DOE
Subject I.D.
123 DRIVE
Operator Name, I.D.
1414 EVERGREEN
Location

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JESSE MARTIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2015

NUMBER 250089

EXPIRES 5/11/2017

MO 580-0721 (5-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

DRUG 580-100

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MARTIN, JESSE
Permit No 250089
Date Issued 5/11/2015 Date Expires 5/11/2017