





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**AARON BAKER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/22/2015

NUMBER 250304

EXPIRES 12/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator BAKER, AARON  
 Permit No 250304  
 Date Issued 12/22/2015 Date Expires 12/22/2017

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 15001**  
**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

AS IV Serial no: 097412  
Version no: 532B

TEST RECORD 00073

Temp Date Time 210L

Air Blank:  
12/24/15 16:35 .000  
Calibration:  
27 12/24/15 16:35 .100

Subject Name

Cal

Subject I.D.

N/A

Operator Name, I.D.

A. Baker 250304

Location

Crocker Police Dept

108 S Commercial

108 S Commercial  
Crocker Police Dept  
A. Baker 250304  
Location  
Operator Name, I.D.  
N/A

AS IV Serial no: 097412  
Version no: 532B  
TEST RECORD 00079  
Temp Date Time 210L  
UOIM: RPY  
12 12/24/15 16:52  
Subject Name  
LFI  
Subject I.D.

108 S Commercial/SR  
Crocker Police Dept  
A. Baker 250304  
Location  
Operator Name, I.D.  
N/A

AS IV Serial no: 097412  
Version no: 532B  
TEST RECORD 00078  
Temp Date Time 210L  
Air Blank:  
12/24/15 16:44 .000  
Subject Test: Man  
29 12/24/15 16:44 .000  
Subject Name  
Blank  
Subject I.D.

AS IV Serial no: 097412  
Version no: 532B

TEST RECORD 00074

Temp Date Time 210L

Air Blank:  
12/24/15 16:36 .000  
Calibration Check:  
28 12/24/15 16:36 .100

Subject Name

Cal check

Subject I.D.

A. Baker 250304

Operator Name, I.D.

A. Baker 250304

Location

Crocker Police Dept

108 S Commercial

108 S Commercial  
Crocker PD  
A. Baker 250304  
Location  
Operator Name, I.D.  
N/A

AS IV Serial no: 097412  
Version no: 532B  
TEST RECORD 00076  
Temp Date Time 210L  
Air Blank:  
12/24/15 16:39 .000  
Calibration Check:  
28 12/24/15 16:39 .099  
Subject Name  
Cal check  
Subject I.D.

108 S Commercial  
Crocker PD  
A. Baker 250304  
Location  
Operator Name, I.D.  
N/A

AS IV Serial no: 097412  
Version no: 532B  
TEST RECORD 00075  
Temp Date Time 210L  
Air Blank:  
12/24/15 16:38 .000  
Calibration Check:  
28 12/24/15 16:38 .100  
Subject Name  
Cal check  
Subject I.D.