



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13060 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on February 7, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.0482% (w/vol) ethyl alcohol. The expiration date for this lot number is February 4, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.040 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN120110-04 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097416
Version no: 004C

TEST RECORD 00247

Temp Date Time ^{s/} 210L

Air Blank:
01/14/15 12:53 .000
Calibration Check:
21 01/14/15 12:53 .041

Subject Name

Subject I.D.

Operator Name, I.D.

DeWayne Carver C41
Location

2920 N. Shamrock Rd.

Jefferson City

AS IV Serial no: 097416
Version no: 004C

TEST RECORD 00248

Temp Date Time ^{s/} 210L

Air Blank:
01/14/15 12:54 .000
Calibration Check:
22 01/14/15 12:54 .041

Subject Name

Subject I.D.

Operator Name, I.D.

DeWayne Carver C41
Location

2920 N. Shamrock Rd.

Jefferson City

AS IV Serial no: 097416
Version no: 004C

TEST RECORD 00249

Temp Date Time ^{s/} 210L

Air Blank:
01/14/15 12:56 .000
Calibration Check:
22 01/14/15 12:56 .040

Subject Name

Subject I.D.

Operator Name, I.D.

DeWayne Carver C41
Location

2920 N. Shamrock Rd.

Jefferson City

AS IV Serial no: 097416
Version no: 004C

TEST RECORD 00250

Temp Date Time ^{s/} 210L

Void: RFI
12 01/14/15 12:57

Subject Name

Subject I.D.

Operator Name, I.D.


DeWayne Carver C41
Location

2920 N. Shamrock Rd.

Jefferson City



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

DEWAYNE D CARVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013
 NUMBER 230101
 EXPIRES 05/28/2015

W. D. Carver

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Vandenberg

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES