

H3

AS IV Serial no: 097442
Version no: 004C

TEST RECORD 00199

Temp Date Time 210L

Air Blank: 01/18/15 20:04 .000
Calibration Check: 23 01/18/15 20:04 .095

Subject Name

Subject I.D.

Operator Name, I.D.

T.A. Hadlock #789

Location

~~MSHP~~ ZONE 4 OFFICE

H2

AS IV Serial no: 097442
Version no: 004C

TEST RECORD 00198

Temp Date Time 210L

Air Blank: 01/18/15 20:03 .000
Calibration Check: 23 01/18/15 20:03 .095

Subject Name

Subject I.D.

Operator Name, I.D.

T.A. Hadlock #789

Location

MSHP ZONE 4 OFFICE

H1

AS IV Serial no: 097442
Version no: 004C

TEST RECORD 00197

Temp Date Time 210L

Air Blank: 01/18/15 20:01 .000
Calibration Check: 22 01/18/15 20:01 .095

Subject Name

Subject I.D.

Operator Name, I.D.

T.A. Hadlock #789

Location

MSHP ZONE 4 OFFICE

AS IV Serial no: 097442
Version no: 004C
TEST RECORD 00200
Temp Date Time 210L
Void: RFI
12 01/18/15 20:06

Subject Name

Subject I.D.

Operator Name, I.D.

T.A. Hadlock #789

Location

MSHP ZONE 4 OFFICE



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE 1 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

**PERMIT
TYPE II**

TODD A HADLOCK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/19/2013

NUMBER 230201

EXPIRES 9/19/2015

MO 580-0771 (6-12)

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named certificate holder is authorized to operate an Alcotest/ breath alcohol instrument for the determination of the alcoholic content in breath from of expired air in Missouri.

OPERATOR: HADLOCK, TODD
Permit No: 230201
Date Issued 9/19/2013 Date Expires 9/19/2015