

21

AS IV Serial no: 099364
Version no: 582B

TEST RECORD 00316
s/
Temp Date Time 210L
Air Blank:
08/24/15 13:57 .000
Calibration Check:
21 08/24/15 13:57 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 582B

TEST RECORD 00318
s/
Temp Date Time 210L
Air Blank:
08/24/15 14:01 .000
Calibration Check:
22 08/24/15 14:01 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 582B

TEST RECORD 00317
s/
Temp Date Time 210L
Air Blank:
08/24/15 13:59 .000
Calibration Check:
22 08/24/15 13:59 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 099364
Version no: 582B

TEST RECORD 00320
s/
Temp Date Time 210L
VOID: NFI
12 08/24/15 14:06

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Aug-2014

Lot # AG423201

Exp. Date

20-Aug-2016

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

391.8 ppm

259.8 ppm

209.0 ppm

103.7 ppm

52.22 ppm

Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

392.5 ppm

258.9 ppm

208.9 ppm

104.9 ppm

52.94 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2014.08.22 12:04:20 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DELBERT A BULLOCK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014
 NUMBER 240129
 EXPIRES 4/3/2016


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
Gail Vesterby
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BULLOCK, DELBERT
 Permit No 240129
 Date Issued 4/3/2014 Date Expires 4/3/2016