



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00293

Temp Date Time ^{9/} 210L

Air Blank:
06/15/15 07:00 .000
Subject Test: Auto
21 06/15/15 07:00 .000

Subject Name

Monthly Maintenance

Subject I.D.

Blank Test

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00294

Temp Date Time ^{9/} 210L

VOID: RFI
12 06/15/15 07:02

Subject Name

Monthly Maintenance

Subject I.D.

RFI Test

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO. 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00295

Temp Date Time ^{9/} 210L

Air Blank:
06/15/15 07:07 .000
Calibration Check:
23 06/15/15 07:07 .096

Subject Name

Monthly Maintenance

Subject I.D.

Test 1

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00296

Temp Date Time ^{9/} 210L

Air Blank:
06/15/15 07:11 .000
Calibration Check:
24 06/15/15 07:11 .097

Subject Name

Monthly Maintenance

Subject I.D.

Test #2

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO 65038

AS IV Serial no: 099365
Version no: 532B

TEST RECORD 00297

Temp Date Time ^{9/} 210L

Air Blank:
06/15/15 07:13 .000
Calibration Check:
24 06/15/15 07:13 .098

Subject Name

Monthly Maintenance

Subject I.D.

Test 3

Operator Name, I.D.

Scott Craig 250048

Location

147 S. Main St.

Laurie, MO 65038



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
SCOTT D CRAIG

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.


DATE 2/20/2015

NUMBER 250048

EXPIRES 2/20/2017

MO 580-0771 (6-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R8-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **CRAIG, SCOTT**
 Permit No **250048**
 Date issued **2/20/2015** Date Expires **2/20/2017**