MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
By Carol Day at 2:04 pm, Aug 25, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
101600	092.3576.581	08/17/2015

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
Midland Empire Alcoholo Task Force Breath Alcohol Test Vehicle	1:45 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☐ SIMULATOR SOLUTION ☑ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Intoximeters LOT # AG402002 EXP. DATE 01/20/2016

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .098	TEST 3 ☛ .099

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	3	1	1	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operating within all limits set forth by the Missouri Department of Health. 0.100 Dry Gas Standard, Lot #AG402002 , Tank 42

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸	Larry R. Stobbs Jr.
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
250134 6/8/2017	(816) 236-1474

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

AS IV Serial no: 101600
Version no: 014D

TEST RECORD 00622

Temp	Date	Time	%BAC
Air Blank:			
	08/17/15	13:45	.000
Subject Test: Man			
19	08/17/15	13:45	.100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 101600
Version no: 014D

TEST RECORD 00623

Temp	Date	Time	%BAC
Air Blank:			
	08/17/15	13:46	.000
Subject Test: Man			
20	08/17/15	13:46	.098

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 101600
Version no: 014D

TEST RECORD 00624

Temp	Date	Time	%BAC
Air Blank:			
	08/17/15	13:48	.000
Subject Test: Man			
20	08/17/15	13:48	.099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 101600
Version no: 014D

TEST RECORD 00625

Temp	Date	Time	%BAC
Void: RFI			
12	08/17/15	13:49	

Subject Name

Subject I.D.

Operator Name, I.D.

Location

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

LARRY R STOBBS JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250134

EXPIRES 6/8/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STOBBS JR, LARRY
Permit No 250134
Date Issued 6/8/2015 **Date Expires** 6/8/2017