



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Brian Lutmer at 1:14 pm, Jan 13, 2015

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>102466</u>	PRINTER SN <u>079 3418 057</u>	DATE OF INSPECTION <u>01/04/2015</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>106 PROGRESS SULLIVAN, MO 63080</u>		TIME OF INSPECTION <u>0630</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 21°C
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION GUTH LABS COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABS LOT # 14200 EXP. DATE 08/05/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34°C SIMULATOR SN SD3322 SIMULATOR EXP DATE 01/28/15

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>.099%</u>	TEST 2 <u>.099%</u>	TEST 3 <u>.100%</u>
---------------------	---------------------	---------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>2</u>	(.15-.19) <u>2</u>	(OVER .19) <u>2</u>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT OPERATING WITHIN P.O.H. SPECIFICATIONS.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME <u>RALPH FINN</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240273 06/13/16</u>	TELEPHONE NUMBER <u>573-468-8001</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 102466
Version no: 532B

TEST RECORD 00109

Temp Date Time 210L s/
VOID: RFI
12 01/04/15 06:30

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 102466
Version no: 532B

TEST RECORD 00110

Temp Date Time 210L s/
Air Blank:
01/04/15 06:32 .000
Calibration Check:
21 01/04/15 06:32 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 102466
Version no: 532B

TEST RECORD 00111

Temp Date Time 210L s/
Air Blank:
01/04/15 06:34 .000
Calibration Check:
21 01/04/15 06:34 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 102466
Version no: 532B

TEST RECORD 00112

Temp Date Time 210L s/
Air Blank:
01/04/15 06:36 .000
Calibration Check:
21 01/04/15 06:36 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

RALPH E FINN JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/13/2014

NUMBER 240273

EXPIRES 6/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator FINN JR, RALPH
Permit No 240273
Date Issued 6/13/2014 Date Expires 6/13/2016