MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

# ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
By Carol Day at 9:11 am, Mar 10, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| ALCO SENSOR IV SN | PRINTER SN | DATE OF INSPECTION |
|---|---|---|
| 102473 | 08C.3527.124 | 02/27/2015 |
| LOCATION OF INSTRUMENT (STREET AND CITY) | | TIME OF INSPECTION |
| 1010 North Boonville Avenue Springfield | | 5:51 pm |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Laboratories LOT # 14220 EXP. DATE 09/24/2016

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34c SIMULATOR SN SD2219 SIMULATOR EXP DATE 01/12/2016

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| TEST 1 ☛ .098 | TEST 2 ☛ .098 | TEST 3 ☛ .098 |
|---|---|---|

☑ RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| REFUSALS | 1 | (0-.04) | 0 | (.05-.09) | 1 | (.10-.14) | 2 | (.15-.19) | 2 | (OVER .19) | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument was found operating within D.O.H. specifications.

**INSPECTING OFFICER**

| SIGNATURE | PRINT NAME |
|---|---|
| ▸ 9/14 | Roger Stewart |
| TYPE II PERMIT NUMBER/EXPIRATION DATE | TELEPHONE NUMBER |
| 250035/ 1-20-2017 | (417) 829-6487 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
[illegible] NORTH 6TH STREET • HARRISBURG, PA 17111-4911 • TELEPHONE: 717-[illegible]

# CERTIFICATE OF ANALYSIS

## Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

***NIST Traceability:***
***Testing was conducted using Cerilliant Reference Standard lot number FN00051301 whose values are traceable to NIST.***
***All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.***

# STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

# PERMIT

## TYPE II

# ROGER A STEWART

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

## ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/20/2015

NUMBER 250035

EXPIRES 1/20/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

**STATE OF MISSOURI**

**DEPARTMENT OF HEALTH AND SENIOR SERVICES**

**BREATH ALCOHOL PROGRAM**

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** STEWART, ROGER

**Permit No** 250035

**Date Issued** 1/20/2015 **Date Expires** 1/20/2017

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00536

| Temp | Date | Time | s/ 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 02/27/15 | 17:51 | .000 |
| Calibration Check: | | | |
| 20 | 02/27/15 | 17:51 | .098 |

Subject Name
TEST 1

Subject I.D.
STEWART, ROGER A

Operator Name, I.D.
250035 / 1-20-2017

Location
LCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00537

| Temp | Date | Time | s/ 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 02/27/15 | 17:53 | .000 |
| Calibration Check: | | | |
| 22 | 02/27/15 | 17:53 | .098 |

Subject Name
TEST 2

Subject I.D.
STEWART, ROGER, A

Operator Name, I.D.
250035 / 1-20-2017

Location
LCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00538

| Temp | Date | Time | s/ 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 02/27/15 | 17:55 | .000 |
| Calibration Check: | | | |
| 23 | 02/27/15 | 17:55 | .098 |

Subject Name
TEST 3

Subject I.D.
STEWART, ROGER, A

Operator Name, I.D.
250035 / 1-20-2017

Location
LCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00539

| Temp | Date | Time | 210L |
|---|---|---|---|

VOID: RFI
12 02/27/15 17:58

Subject Name

RFI

Subject I.D.

STEWART, ROBERT A.

Operator Name, I.D.

250035 / 1-20-2017

Location

LCSO