



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

**RECEIVED**  
 By Carol Day at 11:03 am, Jun 12, 2015

|                             |                            |                                  |
|-----------------------------|----------------------------|----------------------------------|
| ALCO SENSOR IV SN<br>102473 | PRINTER SN<br>08C.3527.124 | DATE OF INSPECTION<br>05/27/2015 |
|-----------------------------|----------------------------|----------------------------------|

|  |                               |
|--|-------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>GREENE COUNTY S/O - 1010 N. BOONVILLE AVE, SPRINGFIELD, MO 65802 | TIME OF INSPECTION<br>8:34 pm |
|--|-------------------------------|

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES LOT # 14220 EXP. DATE 09/24/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2219 SIMULATOR EXP DATE 01/12/2016

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .099

TEST 2  .099

TEST 3  .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |           |   |           |   |           |   |            |   |
|----------|---|---------|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 2 | (0-.04) | (.05-.09) | 1 | (.10-.14) | 6 | (.15-.19) | 1 | (OVER .19) | 1 |
|----------|---|---------|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
RYAN DEVOST

TYPE II PERMIT NUMBER/EXPIRATION DATE  
250077 Expires: 04/29/2017

TELEPHONE NUMBER  
(417) 868-4040

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**RYAN DEVOST**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/29/2015

NUMBER 250077

EXPIRES 4/29/2017

MO 590-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (20-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator DEVOST, RYAN  
Permit No 250077  
Date issued 4/29/2015 Date Expires 4/29/2017

TEST #1

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 00578

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/27/15 20:36 .000  
Calibration Check:  
30 05/27/15 20:36 .099

Subject Name

TEST #1

Subject I.D.

DEVOST, RYAN

Operator Name, I.D.

Permit # 250077 exp: 4/29/17

Location

GCSO

*Ry Devost*

TEST #2

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 00579

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/27/15 20:37 .000  
Calibration Check:  
30 05/27/15 20:37 .099

Subject Name

TEST #2

Subject I.D.

DEVOST, RYAN

Operator Name, I.D.

Permit # 250077 <sup>exp:</sup> #04/29/17

Location

GCSO

*Ry Devost*

TEST #3

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 00580

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/27/15 20:39 .000  
Calibration Check:  
31 05/27/15 20:39 .100

Subject Name

TEST #3

Subject I.D.

Permit # 250077 exp: 04/29/17

Operator Name, I.D.

Devost, RYAN

Location

GCSO

*Ry Devost*

SOBER TEST

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 00577

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
05/27/15 20:34 .000  
Calibration Check:  
30 05/27/15 20:34 .000

Subject Name

SOBER TEST

Subject I.D.

DEVOST, RYAN

Operator Name, I.D.

Permit # 250077 Exp: 4/29/17

Location

GCSO

*Ry Dent*

RFI TEST

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 00581

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 05/27/15 20:47

Subject Name

RFI TEST

Subject I.D.

DEVOST, RYAN

Operator Name, I.D.

Permit # 250077 Exp: 04/29/17

Location

GCSO

*Ry Dent*