



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
REPORT #7
By Carol Day at 1:44 pm, Mar 04, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102488	PRINTER SN 095-3583-748	DATE OF INSPECTION 03/03/2015
LOCATION OF INSTRUMENT (STREET AND CITY) CHARLACK POLICE DEPARTMENT 8401 MIDLAND BLVD. CHARLACK, MO. 63114		TIME OF INSPECTION 2:33 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES LOT # 14200 EXP. DATE 08/05/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2760 SIMULATOR EXP DATE 01/15/2016

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.101 TEST 2 0.101 TEST 3 0.101

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE SGT. <i>[Signature]</i> #64	PRINT NAME SGT. ERIC SONTHEIMER #64
TYPE II PERMIT NUMBER/EXPIRATION DATE 250009 exp 01/02/2017	TELEPHONE NUMBER (314) 427-4715

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-684-8470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IU Serial no: 102468
Version no: 532B

TEST RECORD 00175

Temp Date Time ^{9/} 210L

Air Blank:
03/03/15 14:33 .000
Calibration Check:
24 03/03/15 14:33 .101

Subject Name

SIMULATOR

Subject I.D.

Operator Name, I.D.

SGT — GS

Location

CPD

AS IU Serial no: 102468
Version no: 532B

TEST RECORD 00176

Temp Date Time ^{9/} 210L

Air Blank:
03/03/15 14:37 .000
Calibration Check:
25 03/03/15 14:37 .101

Subject Name

SIMULATOR

Subject I.D.

Operator Name, I.D.

SGT — 07

Location

CPD

AS IU Serial no: 102468
Version no: 532B

TEST RECORD 00177

Temp Date Time ^{9/} 210L

Air Blank:
03/03/15 14:39 .000
Calibration Check:
25 03/03/15 14:39 .101

Subject Name

SIMULATOR

Subject I.D.

Operator Name, I.D.

SGT — GS

Location

CPD

AS IU Serial no: 102468
Version no: 532B

TEST RECORD 00179

Temp Date Time ^{9/} 210L

VOID: RFI
12 03/03/15 14:43

Subject Name

SIMULATOR

Subject I.D.

Operator Name, I.D.

SGT — GS

Location

CPD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

ERIC C SONTHEIMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 1/2/2015

NUMBER 250009

EXPIRES 1/2/2017

MO 560-0771 (8-10)



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **SONTHEIMER, ERIC**
Permit No **250009**
Date Issued **1/2/2015** Date Expires **1/2/2017**