RECEIVED
By Carol Day at 2:33 pm, May 18, 2015

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
104293	09B.3590.034	5-6-15
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
7010 N HOL MES		0400

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [ ] SIMULATOR SOLUTION
- [x] COMPRESSED ETHANOL-GAS MIXTURE
- [ ] STANDARD SUPPLIER INTOXIMETERS LOT # A6421804 EXP. DATE 08-06-16
- [ ] SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____
- [ ] CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [ ] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [x] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.082	.081	.081

- [x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
1	0	0	1	2	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Stan Troncin	STAN TRONCIN
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240402 11/13/16	816-436-3550

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

MO 580-1351 (6-10) AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER LAB 114
services provided on a nondiscriminatory basis

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo 63103
Ph (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 6-Aug-2014

Lot # AG421804

Exp. Date	Cyl. Type	Component	Certified Concentration
6-Aug-2016	108	Ethanol	0.080 ± 0.002 BrAC (218 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2014.08.06 17:38:54 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 104293
Version no: 532B

TEST RECORD 00276

Temp Date Time 210L g/

Air Blank:
05/06/15 07:28 .000
Calibration Check:
23 05/06/15 07:28 .082

Subject Name

Subject I.D.
TRONCIN 16956
Operator Name, I.D.

Location

AS IV Serial no: 104293
Version no: 532B

TEST RECORD 00277

Temp Date Time 210L g/

Air Blank:
05/06/15 07:30 .000
Calibration Check:
24 05/06/15 07:30 .081

Subject Name

Subject I.D.
TRONCIN 16956
Operator Name, I.D.

Location

AS IV Serial no: 104293
Version no: 532B

TEST RECORD 00278

Temp Date Time 210L g/

Air Blank:
05/06/15 07:32 .000
Calibration Check:
25 05/06/15 07:32 .081

Subject Name

Subject I.D.
TRONCIN 16956
Operator Name, I.D.

Location

AS IV Serial no: 104293
Version no: 532B

TEST RECORD 00279

Temp Date Time 210L g/

VOID: RFI
12 05/06/15 07:34

Subject Name

Subject I.D.
TRONCIN 16956
Operator Name, I.D.

Location

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

STANLEY F TRONCIN JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306 111 through 306 119 RSMo.

DATE 11/13/2014

NUMBER 240402

EXPIRES 11/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6 10)

MO 580-0771 (6 12)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator TRONCIN JR STANLEY
Permit No 240402
Date Issued 11/13/2014 Date Expires 11/13/2016