



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107977	PRINTER SN 096.3580.877	DATE OF INSPECTION 11/02/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 301 S. Main St, Excelsior Springs, MO		TIME OF INSPECTION 8:45 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LOT # 15120 EXP. DATE 04/29/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN DR5383 SIMULATOR EXP DATE 06/05/2016

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 $\blacktriangleleft$ .096	TEST 2 $\blacktriangleleft$ .097	TEST 3 $\blacktriangleleft$ .097
----------------------------------	----------------------------------	----------------------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	0	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Time changed due to daylight savings.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Brian K Kennedy
TYPE II PERMIT NUMBER/EXPIRATION DATE 250071 / 03/31/2017	TELEPHONE NUMBER (816) 630-2000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 187977  
Version no: 532B

TEST RECORD 00176

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/02/15 08:46 .000  
Calibration Check:  
23 11/02/15 08:46 .096

Subject Name

MONTHLY MAINT  
Subject I.D.

Operator Name, I.D.

KENNEY 911

Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 187977  
Version no: 532B

TEST RECORD 00177

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/02/15 08:48 .000  
Calibration Check:  
23 11/02/15 08:48 .097

Subject Name

SIMPLER  
Subject I.D.

Operator Name, I.D.

KENNEY 911

Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 187977  
Version no: 532B

TEST RECORD 00178

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/02/15 08:50 .000  
Calibration Check:  
23 11/02/15 08:50 .097

Subject Name

MONTHLY MAINT  
Subject I.D.

Operator Name, I.D.

KENNEY 911

Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 187977  
Version no: 532B

TEST RECORD 00179

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 11/02/15 08:51

Subject Name

MONTHLY MAINT  
Subject I.D.

Operator Name, I.D.

KENNEY 911

Location

301 S MAIN ST

EX SPG, MO

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**BRIAN K KENNEDY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **3/31/2015**

NUMBER **250071**

EXPIRES **3/31/2017**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-3-92-10

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **KENNEDY, BRIAN**  
Permit No **250071**  
Date Issued **3/31/2015** Date Expires **3/31/2017**