MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 10:48 am, May 04, 2015

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
107981	099.3586.798	05/01/2015

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
1 Court Circle Suite 13, Camdenton,MO,65020	12:20 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [ ] SIMULATOR SOLUTION
- [x] COMPRESSED ETHANOL-GAS MIXTURE

- [ ] STANDARD SUPPLIER Intoximeters LOT # AG505101 EXP. DATE 02/20/2017

- [ ] SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .100	TEST 3 ☛ .100

- [x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	2

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibration needed, and completed.

INSPECTING OFFICER

SIGNATURE	PRINT NAME
	Jason A. Sylvester
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
280346 / 12/30/2015	(573) 346-2243

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JASON A SYLVESTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/31/2013

NUMBER 230346

EXPIRES 12/31/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling ,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SYLVESTER, JASON
Permit No 230346
Date Issued 12/31/2013 **Date Expires** 12/31/2015

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 23-Feb-2015

Lot # AG505101

Exp. Date	Cyl. Type	Component	Certified Concentration
20-Feb-2017	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.02.23 15:38:13 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 107981
Version no: 532B

TEST RECORD 00039

Temp	Date	Time	9/ 210L
Air Blank:			
	05/01/15	12:19	.000
Calibration:			
23	05/01/15	12:19	.100

Subject Name

Subject I.D.

Operator Name, I.D.
230346
J. Sylvester 24602

Location
C.C.S.O

AS IV Serial no: 107981
Version no: 532B

TEST RECORD 00040

Temp	Date	Time	9/ 210L
Air Blank:			
	05/01/15	12:20	.000
Calibration Check:			
23	05/01/15	12:20	.100

Subject Name

Subject I.D.

Operator Name, I.D.
230346
J. Sylvester 24602

Location
C.C.S.O

AS IV Serial no: 107981
Version no: 532B

TEST RECORD 00041

Temp	Date	Time	9/ 210L
Air Blank:			
	05/01/15	12:22	.000
Calibration Check:			
24	05/01/15	12:22	.100

Subject Name

Subject I.D.

Operator Name, I.D.
230346
J. Sylvester 24602

Location
C.C.S.O

AS IV Serial no: 107981
Version no: 532B

TEST RECORD 00042

Temp	Date	Time	9/ 210L
Air Blank:			
	05/01/15	12:23	.000
Calibration Check:			
25	05/01/15	12:23	.100

Subject Name

Subject I.D.

Operator Name, I.D.
230346
J. Sylvester 24602

Location
C.C.S.O

AS IV Serial no: 107981
Version no: 532B

TEST RECORD 00043

Temp	Date	Time	9/ 210L
VOID: RFI			
12	05/01/15	12:28	

Subject Name

Subject I.D.

Operator Name, I.D.
230346
J. Sylvester 24602

Location
C.C.S.O