



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file. *By Carol Day at 8:27 am, Aug 10, 2015 red.*

ALCO SENSOR IV SN 107989	PRINTER SN 099.3586.794	DATE OF INSPECTION 08/04/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 1700 US Hwy 67 Florissant Missouri 63031		TIME OF INSPECTION 6:10 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERES LOT # AG332301 EXP. DATE 11/19/2015
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <i>☛</i> .098	TEST 2 <i>☛</i> .098	TEST 3 <i>☛</i> .097
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument working within D.O.H. guidelines.

INSPECTING OFFICER

SIGNATURE <i>Daniel Howard</i>	PRINT NAME Howard, Daniel 575
TYPE II PERMIT NUMBER/EXPIRATION DATE 250052 / 02/20/2017	TELEPHONE NUMBER (314) 837-7000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 20-Nov-2013

Lot # AG332301

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
19-Nov-2015	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2013.11.20 15:33:36 -06:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IU Serial no: 107989
Version no: 532B

TEST RECORD 00052
Temp Date Time 210L

Air Blank: 08/04/15 06:10 .000
Calibration Check: 23 08/04/15 06:10 .098

Subject Name
Maintenance
Subject I.D.
575

Operator Name, I.D.
Howard 575
Location
1700 N. Hwy 67

Plantment, mo 08007

AS IU Serial no: 107989
Version no: 532B

TEST RECORD 00053
Temp Date Time 210L

Air Blank: 08/04/15 06:12 .000
Calibration Check: 24 08/04/15 06:12 .098

Subject Name
Maintenance
Subject I.D.
575

Operator Name, I.D.
Howard 575
Location
1700 N. Hwy 67

Plantment, mo 08007

AS IU Serial no: 107989
Version no: 532B

TEST RECORD 00054
Temp Date Time 210L

Air Blank: 08/04/15 06:13 .000
Calibration Check: 25 08/04/15 06:13 .097

Subject Name
Maintenance
Subject I.D.
575

Operator Name, I.D.
Howard 575
Location
1700 N. Hwy 67

Plantment, mo 08007

AS IU Serial no: 107989
Version no: 532B

TEST RECORD 00055
Temp Date Time 210L

VOID: RFI
12 08/04/15 06:14

Subject Name
Maintenance
Subject I.D.
575

Operator Name, I.D.
Howard 575
Location
1700 N. Hwy 67

Plantment, mo 08007

AS IU Serial no: 107989
Version no: 532B
TEST RECORD 00056
Temp Date Time 210L
Air Blank: 08/04/15 06:16 .000
Subject Test: Auto 25 08/04/15 06:16 .000
Subject Name
Maintenance
Subject I.D.
575
Operator Name, I.D.
Howard 575
Location
1700 N. Hwy 67
Plantment, mo 08007



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DANIEL HOWARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of section 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2015

NUMBER 250052

EXPIRES 2/20/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-0771 (8-10)

LAB-4 (R6-1)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HOWARD, DANIEL
 Permit No 250052
 Date Issued 2/20/2015 Date Expires 2/20/2017