



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 11:17 am, May 05, 2015

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108264	PRINTER SN 099.3586.827	DATE OF INSPECTION 05/02/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 315 S. Hudson, Buckner MO		TIME OF INSPECTION 4:10 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .101

TEST 2 → .101

TEST 3 → .100

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	4	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE ▶	PRINT NAME
TYPE II PERMIT NUMBER/EXPIRATION DATE 240421 11-24-2016	TELEPHONE NUMBER (816) 650-3939

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

600 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-554-6420

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 188264
Version no: 532B
TEST RECORD 00064
Temp Date Time 210L 9/
Air Blank: 05/02/15 04:27 .000
Subject Test: Man 25 05/02/15 04:27 .101
Subject Name
Subject I.D.
Operator Name, I.D.
Location
BPP
Mantione

AS IV Serial no: 188264
Version no: 532B
TEST RECORD 00065
Temp Date Time 210L 9/
Air Blank: 05/02/15 04:34 .000
Subject Test: Auto 27 05/02/15 04:34 .000
Subject Name
Subject I.D.
Operator Name, I.D.
Location
BPP
Mantione

AS IV Serial no: 188
Version no: 532B
TEST RECORD 00067
Temp Date Time 210L 9/
Air Blank: 05/02/15 04:31 .000
Subject Test: Auto 26 05/02/15 04:31 .100
Subject Name
Subject I.D.
Operator Name, I.D.
Location
BPP
Mantione

AS IV Serial no: 188264
Version no: 532B
TEST RECORD 00078
Temp Date Time 210L 9/
VOID: RFI 12 05/02/15 04:37
Subject Name
Subject I.D.
Operator Name, I.D.
Location
BPP
Mantione



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

LAUREN R ONKA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/24/2014

NUMBER 240421

EXPIRES 11/24/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 593-0771 (6-10)

LAG-4 (716-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ONKA, LAUREN
Permit No 240421
Date Issued 11/24/2014 Date Expires 11/24/2016