



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
By Carol Day at 1:04 pm, Jul 13, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108265	PRINTER SN 099.3586.806	DATE OF INSPECTION 06/29/2015
LOCATION OF INSTRUMENT (STREET AND CITY) Elsberry Police Department		TIME OF INSPECTION 2:17 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Lab LOT # 14110 EXP. DATE 05/10/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2287 SIMULATOR EXP DATE 06/02/2016

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ~~0.095~~ .096

TEST 2 ~~0.096~~ .096

TEST 3 ~~0.095~~ .095

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.16-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

07/09/2015 took into intoxmeter to have button release fixed

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Cpl. David Wait

TYPE & PERMIT NUMBER/EXPIRATION DATE
240416

TELEPHONE NUMBER
(573) 898-5456

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

640 NORTH 87th STREET • HARRISBURG, PA 17111-4311 • TELEPHONE: 717-561-4470

CERTIFICATE OF ANALYSIS**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Certified Reference Standard lot number EN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 108265
 Version no: 532B

TEST RECORD 00146

Temp Date Time 210L

Air Blank:
 07/29/15 02:15 .000

Calibration Check:
 22 07/23/15 02:15 .095

Subject Name
 Test 1

Subject I.D.

Operator Name, I.D.
 DL Wait

Location
 Elsberry P.D.

AS IV Serial no: 108265
 Version no: 532B

TEST RECORD 00148

Temp Date Time 210L

Air Blank:
 07/29/15 02:19 .000

Calibration Check:
 22 07/23/15 02:19 .095

Subject Name
 Test 1

Subject I.D.

Operator Name, I.D.
 DL Wait

Location
 Elsberry P.D.

AS IV Serial no: 108265
 Version no: 532B

TEST RECORD 00148

Temp Date Time 210L

Air Blank:
 07/29/15 02:18 .000

Calibration Check:
 22 07/23/15 02:18 .095

Subject Name
 TEST 2

Subject I.D.

Operator Name, I.D.
 Cpl. Wait

Location
 Elsberry P.D.

AS IV Serial no: 108265
 Version no: 532B

TEST RECORD 00147

Temp Date Time 210L

Air Blank:
 07/29/15 02:16 .000

Calibration Check:
 21 07/29/15 02:16 .096

Subject Name
 Test 3

Subject I.D.

Operator Name, I.D.
 Cpl. Wait

Location
 Elsberry P.D.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DAVID M WAIT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/20/2014

NUMBER 240416

EXPIRES 11/20/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R8-10)

MO 689-0771 (8-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WAIT, DAVID
Permit No 240416
Date Issued 11/20/2014 Date Expires 11/20/2016