



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

REPORT #7

By Carol Day at 4:16 pm, Jun 29, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108268	PRINTER SN 099.3586.791	DATE OF INSPECTION 6/25/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 1781 Zumbetti Rd, St Charles, MO 63303		TIME OF INSPECTION 1825hr

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) **ok**
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **ok**
- PRINTER WORKING PROPERLY **ok**
- TIME AND DATE DISPLAYING PROPERLY **ok**

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL GAS MIXTURE
- STANDARD SUPPLIER **Guth Laboratories** LOT # **14030** EXP. DATE **1/20/2016**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD2248** SIMULATOR EXP DATE **1/15/2016**
- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .100	TEST 3 .100
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING **ok**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 2	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE Daniel Gibbons 324	PRINT NAME Daniel Gibbons 324
TYPE II PERMIT NUMBER/EXPIRATION DATE 250021 1/20/2017	TELEPHONE NUMBER (636) 949-3300 x2324

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 108268
Version no: 532B

TEST RECORD 00177

Temp Date Time ^{s/} 210L

Air Blank:
06/25/15 18:25 .000
Calibration Check:
21 06/25/15 18:25 .101

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbethl Rd.

St. Charles, mo 63303

AS IV Serial no: 108268
Version no: 532B

TEST RECORD 00178

Temp Date Time ^{s/} 210L

Air Blank:
06/25/15 18:26 .000
Calibration Check:
22 06/25/15 18:26 .100

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbethl Rd.

St. Charles, MO 63303

AS IV Serial no: 108268
Version no: 532B

TEST RECORD 00179

Temp Date Time ^{s/} 210L

Air Blank:
06/25/15 18:28 .000
Calibration Check:
22 06/25/15 18:28 .100

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbethl Rd.

St. Charles, Mo 63303

AS IV Serial no: 108268
Version no: 532B

TEST RECORD 00180

Temp Date Time ^{s/} 210L

VOID: RFI
12 06/25/15 18:29

Subject Name

Subject I.D.

Operator Name, I.D.

1781 Zumbethl Rd.

Location

St. Charles, Mo 63303

D. Gibbons 324