

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00069

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/23/15 14:39 .000
Calibration:
23 11/23/15 14:39 .100

Subject Name

Calibration

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City, MO

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00070

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/23/15 14:43 .000
Calibration Check:
24 11/23/15 14:43 .100

Subject Name

MAINT Check

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City, MO

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00071

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/23/15 14:44 .000
Calibration Check:
25 11/23/15 14:44 .101

Subject Name

MAINT Check

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City, MO

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00072

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/23/15 14:47 .000
Calibration Check:
25 11/23/15 14:47 .100

Subject Name

MAINT Check

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City, MO

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00073

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 11/23/15 14:48

Subject Name

RFI-Test

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland

Location

2920 N. Shamrock

Jefferson City, MO



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

JIMMY L CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250155

EXPIRES 7/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator CLEVELAND, JIMMY
 Permit No 250155
 Date Issued 7/22/2015 Date Expires 7/22/2017

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.