

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00083

Temp Date Time ^{a/} 210L

Air Blank:
08/27/15 13:21 .000
Calibration Check:
26 08/27/15 13:21 .099

Subject Name

Subject I.D.

Operator Name, I.D.

TPR. J.L. WHITE 1202

Location

TROOP D ZONE 3

ZONE OFFICE

BOLZUAR

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00081

Temp Date Time ^{a/} 210L

Air Blank:
08/27/15 13:18 .000
Calibration Check:
25 08/27/15 13:18 .100

Subject Name

Subject I.D.

Operator Name, I.D.

TPR. J.L. WHITE 1202

Location

TROOP D ZONE 3

ZONE OFFICE

BOLZUAR

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00084

Temp Date Time ^{a/} 210L

VOID: RFI
12 08/27/15 13:22

Subject Name

Subject I.D.

Operator Name, I.D.

TPR. J.L. WHITE 1202

Location

TROOP D ZONE 3

ZONE OFFICE

BOLZUAR

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00082

Temp Date Time ^{a/} 210L

Air Blank:
08/27/15 13:19 .000
Calibration Check:
26 08/27/15 13:19 .099

Subject Name

Subject I.D.

Operator Name, I.D.

TPR. J.L. WHITE 1202

Location

TROOP D ZONE 3

ZONE OFFICE

BOLZUAR



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JOSHUA L WHITE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240080

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **WHITE, JOSHUA**
 Permit No **240080**
 Date Issued **3/7/2014** Date Expires **3/7/2016**



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Gail Vasterling
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Missouri State Highway Patrol
 Serial Number: MP2146
 Manufacturer: Guth
 Model Number: 12V500

CALIBRATION RESULTS

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
34.00	34.00

This calibration was performed with NIST-Traceable Thermometer SN: 307715

This calibration was performed by: Jim Cleveland

This calibration was performed: 06/09/15

COPY OF CALIBRATION STICKER

Serial Num	<u>MP2146</u>	By	<u>JLC</u>
Ref Temp	<u>34.00</u> °C	Sim Temp	<u>34.00</u> °C
Certification Expires	<u>6/9/16</u>		
Ref Therm Ser#	<u>307715</u>	Date	<u>6/9/15</u>