

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00141

Temp Date Time ^{g/} 210L

Air Blank:
08/31/15 08:25 .000
Calibration Check:
24 08/31/15 08:25 .092

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

J. Snodgrass 440

Location

Cottleville P.D.

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00140

Temp Date Time ^{g/} 210L

Air Blank:
08/31/15 08:24 .000
Calibration Check:
23 08/31/15 08:24 .092

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

J. Snodgrass 440

Location

Cottleville P.D.

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00139

Temp Date Time ^{g/} 210L

Air Blank:
08/31/15 08:22 .000
Calibration Check:
22 08/31/15 08:22 .092

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

J. Snodgrass 440

Location

Cottleville P.D.

Cottleville P.D.

Operator Name, I.D.
J. Snodgrass 440
Location

Subject Name
RFI Test
Subject I.D.

UQID: RFI
12 08/31/15 08:28
Temp Date Time ^{g/} 210L

AS IU Serial no: 108392
Version no: 532B
TEST RECORD 00142

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-684-6470

CERTIFICATE OF ANALYSIS**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19 CSR 25-30.051 (4).

Client: 03802015 Expires: 03/31/2016
Dulac Thermal, SNC 024702
MSC Test, DRL Temp: 34.00
Agency: Cordwaine Police Dept
SD 2588
MISSOURI SAFETY CENTER
MISSOURI SAFETY CENTER

Technician Printed Name: DAN LUCAS

Technician Signature: [Handwritten Signature]

Date: 05/30/2015

Contact: Missouri Safety Center
Breath Alcohol Instrument Training Program

2

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

JAMIE E SNODGRASS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the calibration of the alcoholic content of blood from a sample of exhaled air. Permit issued under the provisions of sections 577.09D through 577.041, RSMo and 306.111 through 306.119, RSMo.

DATE 3/25/2015
NUMBER ZSUD67
EXPIRES 3/25/2017
15-28-0718-10

Wm. M. ...
DIRECTOR OF MISSOURI PUBLIC HEALTH LABORATORY
Shel Vachek
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD
Operator: JAMIE E SNODGRASS
Permit No: ZSUD67
Plate: 15-28-0718-10
15-28-0718-10