



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**

REPORT #6

By Carol Day at 2:06 pm, Aug 31, 2015

Complete this report at the time of the regular monthly preventive maintenance of the instrument.  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 2000591	NAME OF AGENCY Directorate of Emergency Services	DATE OF INSPECTION 08/31/2015
LOCATION OF INSTRUMENT (STREET AND CITY) Bldg 1000, Fort Leonard Wood		TIME OF INSPECTION 0657

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 08/31/15 0657
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GUTH LABS INC LOT # 14220 EXP. DATE 09/24/2016	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5376 EXP. DATE 10/20/2015	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .096	TEST 2 • .098	TEST 3 • .098
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0	(0-.04) 1	(.05-.09) 1	(.10-.14) 2	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

TESTED AND CERTIFIED AS WITHIN GUIDELINES ESTABLISHED BY DHSS.

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME ROBERT ISHMAEL
TYPE II PERMIT NUMBER/EXPIRATION DATE 250005 01/02/2017	TELEPHONE NUMBER (573) 596-1074

**RETURN COMPLETED REPORT TO THE:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

FORT LEONARD WOOD  
DIA SUPPORT & SERVICE CENTER

NBC DATA MASTER SERIAL NUMBER 000010  
28/01/73  
28:57


--- INSTRUMENTS CHECK ---

COMPUTER:	OKAY
PROGRAM (000-67-2009):	OKAY
WEIGHTS	
SAMPLE CHAMBER:	OK
FLOW DETECTOR:	OKAY
FLY	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARDS:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#\$%^&\*()'+,-./0123456789:;|<=>?[]`~  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz  
pqrstuvwxyz{|}~

OPERATOR SIGNATURE



Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.  
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

# BAC DataMaster

Evidence Ticket

FORT LEONARD WOOD  
DIRECTORATE OF EMERGENCY SERVICES

BAC DATA MASTER SERIAL NUMBER 200059  
08/31/15

TESTING OFFICER:  
ISRAEL ROBERT  
OFFICER I.D.: 18748  
PERMIT NUMBER: 250005  
EXPIRATION DATE: 01/02/17  
MISCELLANEOUS DATA:

----- SUPERVISOR NAME -----

BLANK TEST	.000	07:14
INTERNAL STANDARD	VERIFIED	07:14
EXTERNAL STANDARD	.056	07:14
BLANK TEST	.000	07:15
EXTERNAL STANDARD	.098	07:15
BLANK TEST	.000	07:16
EXTERNAL STANDARD	.098	07:16
BLANK TEST	.000	07:17

n = 3  
SIM. = .1  
AVG. = .0973

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

# BAC DataMaster

Evidence Ticket

FORT LEONARD WOOD  
DIRECTORATE OF EMERGENCY SERVICES

BAC DATA MASTER SERIAL NUMBER 200059  
08/31/15

ARREST TIME: 01:00  
SUBJECT NAME:  
NA  
JOB: 01/01/01 SEX: M  
STATE/D.L.: NA/NA  
ARRESTING OFFICER:  
NA

OFFICER I.D.: NA  
TESTING OFFICER:  
ISRAEL ROBERT  
OFFICER I.D.: 18748  
PERMIT NUMBER: 250005  
EXPIRATION DATE: 01/02/17  
MISCELLANEOUS DATA:  
RFL/TEST

----- BREATH ANALYSIS -----

BLANK TEST	.000	07:22
INTERNAL STANDARD	VERIFIED	07:22
RADIO INTERFERENCE		



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**ROBERT A ISHMAEL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/2/2015

NUMBER 250005

EXPIRES 1/2/2017

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
 ,acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator ISHMAEL, ROBERT  
 Permit No 250005  
 Date Issued 1/2/2015 Date Expires 1/2/2017



# GUTH LABORATORIES, INC.

590 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL SAN-610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN00051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.