



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>201203</b>	NAME OF AGENCY <b>Chaffee Police Department</b>	DATE OF INSPECTION <b>03-28-2015</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>222 W. Yoakum Ave Chaffee</b>		TIME OF INSPECTION <b>15:22 hours</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>03-28-2015 @ 15:22 hrs</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>+ 49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <b>Repco Marketing</b> LOT # <b>13002</b> EXP. DATE <b>6-19-2015</b>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <b>+ 34.0</b> °C SIMULATOR SN <b>5D2228</b> EXP. DATE <b>3-12-2016</b>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <b>.098</b>	TEST 2 <b>.098</b>	TEST 3 <b>.099</b>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>0</b>	(0-.04) <b>0</b>	(.05-.09) <b>0</b>	(.10-.14) <b>1</b>	(.15-.19) <b>1</b>	OVER .19 <b>0</b>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).  
**None.**

**INSPECTING OFFICER**

SIGNATURE <b>William Sammut</b>	PRINT FULL NAME <b>William Sammut</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>250014 1-14-2017</b>	TELEPHONE NUMBER <b>573-887-6911</b>

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

STATE OF MISSOURI  
CHAFFEE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201203  
03/28/15  
15:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 49c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abdefghijklmno  
pqrstuvwxyz{|}~

Operator Signature

*W. Jammal* SEMO  
DPS 521

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

STATE OF MISSOURI  
CHAFFEE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201203  
03/28/15

TESTING OFFICER:  
SAMMUT/WILLIAM/J  
OFFICER I.D.: SEMO 521  
PERMIT NUMBER: 250014  
EXPIRATION DATE: 01/14/17  
MISCELLANEOUS DATA:  
MARCH MAINT REPORT

--- SUPERVISOR MODE ---

BLANK TEST	.000	16:15
INTERNAL STANDARD	VERIFIED	16:15
EXTERNAL STANDARD	.099	16:15
BLANK TEST	.000	16:16
EXTERNAL STANDARD	.099	16:16
BLANK TEST	.000	16:17
EXTERNAL STANDARD	.099	16:17
BLANK TEST	.000	16:18

N = 3  
SIM. = .1  
AVG. = .0983

Operator Signature

*W. Sammut* SEMO 521  
DPS

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
CHAFFEE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201203  
03/28/15

ARREST TIME: 15:00  
SUBJECT NAME:  
SANNUT/WILLIAM/J  
DOB: 01/01/01 SEX: M  
STATE/D.L.: MO/123456789  
ARRESTING OFFICER:  
SANNUT/WILLIAM/J  
OFFICER I.D.: SEMO 521  
TESTING OFFICER:  
SANNUT/WILLIAM/J  
OFFICER I.D.: SEMO 521  
PERMIT NUMBER: 250014  
EXPIRATION DATE: 01/14/17  
MISCELLANEOUS DATA:  
MARCH MAINT REPORT  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 16:23  
INTERNAL STANDARD VERIFIED 16:23  
RADIO INTERFERENCE

Operator Signature \_\_\_\_\_

*W. Sannut* SEMO  
DPS 521



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**

**WILLIAM J SAMMUT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/14/2015

*William J Sammut*

NUMBER 250014

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

EXPIRES 1/14/2017

*Paul Verkley*

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.