



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH-ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 2:30 pm, Apr 29, 2015

REPORT

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>201222</u>	NAME OF AGENCY <u>Dixon Police Dept</u>	DATE OF INSPECTION <u>4/20/15</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>503 South Elm St Dixon Mo 65459</u>		TIME OF INSPECTION <u>0930</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed value where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>4/20/15 0955</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Guth Labs</u>	LOT # <u>14220</u> EXP. DATE <u>9/24/16</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34</u> °C	SIMULATOR SN <u>SD2279</u> EXP. DATE <u>12/3/15</u>
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • <u>.094</u>	TEST 2 • <u>.096</u>	TEST 3 • <u>.097</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) OK

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	OVER .19 <u>0</u>
----------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY): working within Doff spec

INSPECTING OFFICER	
SIGNATURE <u>[Signature]</u>	PRINT FULL NAME <u>Mike Plummer</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>250007</u>	TELEPHONE NUMBER <u>573-759-6610</u>
RETURN COMPLETED REPORT TO THE:	<u>01/02/17</u>

Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 87th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**CERTIFIED ALCOHOL REFERENCE
SOLUTION FOR SIMULATOR**

14220 9/24/14 9/24/16
LOT NO. MFG. DATE EXP. DATE

275 Gal. 500 ML _____
LOT VOL. BOT. VOL. BOT. NO.

When this reference solution is used with a breath simulator certified by Guth Laboratories, a properly operating instrument will read 0.10
For additional information contact:

Guth Laboratories, Inc.
590 North 67th Street, Harrisburg, PA 17111
Toll Free 800-233-2338
Rev. 4/02



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

DATE OF ANALYSIS

ANALYST'S NAME AND TITLE (PRINT OR TYPE)
DATE

IDENTIFY THE ALCOHOL
CONCENTRATION
OFFICER'S NAME AND TITLE
PERMIT NUMBER (if applicable)
EXPIRATION DATE (if applicable)
MISCELLANEOUS NOTES

OFFICER'S ID NO.

ALCOHOL TEST	DATE	TIME
PRELIMINARY SCREENING	MM/DD/YY	HH:MM
EXTENDED SCREENING	MM/DD/YY	HH:MM
IR FINGER TEST	MM/DD/YY	HH:MM
EXTENDED SCREENING	MM/DD/YY	HH:MM
IR FINGER TEST	MM/DD/YY	HH:MM
EXTENDED SCREENING	MM/DD/YY	HH:MM
IR FINGER TEST	MM/DD/YY	HH:MM

REMARKS
SAMPLE NO.
ANALYST'S SIGNATURE

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

DATE OF ANALYSIS

ANALYST'S NAME AND TITLE (PRINT OR TYPE)
DATE

IDENTIFY THE ALCOHOL
CONCENTRATION
OFFICER'S NAME AND TITLE
PERMIT NUMBER (if applicable)
EXPIRATION DATE (if applicable)
MISCELLANEOUS NOTES

REMARKS
SAMPLE NO.
ANALYST'S SIGNATURE

Operator Signature _____

2208-02

Operator Signature _____

2208-02

